

LB125184

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

4/27/2023 13:06

Reviewed by : 12 Instrument ID : Konelab

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.975	0.0	0.112	
ICB1	-0.002	0.0	0.007	
CCV1	0.944	0.0	0.109	
CCB1	0.002	0.0	0.008	
PB152445BL	-0.001	0.0	0.008	
PB152445BS	0.973	0.0	0.112	
O2468-01	0.049	0.0	0.013	
O2495-01	-0.742	0.0	-0.072	
O2495-03	-0.400	0.0	-0.035	
O2495-05	0.019	0.0	0.010	
O2495-07	0.021	0.0	0.010	
O2495-09	-0.015	0.0	0.006	
O2521-01	-0.958	0.0	-0.095	
O2521-03	-1.355	0.0	-0.138	
CCV2	0.933	0.0	0.108	
CCB2	-0.004	0.0	0.007	
O2521-05	-0.721	0.0	-0.070	
O2521-07	-0.564	0.0	-0.053	
O2521-07DUP	-0.558	0.0	-0.052	
O2521-09MS	0.443	0.0	0.055	
12 O2521-10MSD	0.411	0.0	0.052	
O2521-14	-0.002	0.0	0.007	
CCV3	0.945	0.0	0.109	
CCB3	-0.003	0.0	0.007	

N 24
Mean 0.016
SD 0.6436
CV% 3971.09

Aquakem v. 7.2AQ1

Results from time period:

Thu Apr 27 10:47:25 2023

Thu Apr 27 12:24:11 2023

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0033	mg/l	4/27/2023 10:47:25	
0.05PPM	A	Phenolics-	P	0.0498	mg/l	4/27/2023 10:47:26	
0.1PPM	A	Phenolics-	P	0.0975	mg/l	4/27/2023 10:47:27	
0.25PPM	A	Phenolics-	P	0.2464	mg/l	4/27/2023 10:47:28	
0.50PPM	A	Phenolics-	P	0.4998	mg/l	4/27/2023 10:47:29	
1.0PPM	A	Phenolics-	P	1.0051	mg/l	4/27/2023 10:47:30	
2.0PPM	A	Phenolics-	P	1.9981	mg/l	4/27/2023 10:47:31	
ICV1	S	Phenolics-	P	0.9751	mg/l	4/27/2023 11:36:19	
ICB1	S	Phenolics-	P	-0.0024	mg/l	4/27/2023 11:36:22	
CCV1	S	Phenolics-	P	0.9437	mg/l	4/27/2023 11:36:24	
CCB1	S	Phenolics-	P	0.0023	mg/l	4/27/2023 11:36:26	
PB152445BL	S	Phenolics-	P	-0.0009	mg/l	4/27/2023 11:36:27	
PB152445BS	S	Phenolics-	P	0.9728	mg/l	4/27/2023 11:36:30	
O2468-01	S	Phenolics-	P	0.0488	mg/l	4/27/2023 11:46:21	
O2495-01	S	Phenolics-	P	-0.7424	mg/l	4/27/2023 11:46:22	
O2495-03	S	Phenolics-	P	-0.3997	mg/l	4/27/2023 11:46:23	
O2495-05	S	Phenolics-	P	0.0188	mg/l	4/27/2023 11:46:24	
O2495-07	S	Phenolics-	P	0.0207	mg/l	4/27/2023 11:46:25	
O2495-09	S	Phenolics-	P	-0.0146	mg/l	4/27/2023 11:46:26	
O2521-01	S	Phenolics-	P	-0.9585	mg/l	4/27/2023 11:46:27	
O2521-03	S	Phenolics-	P	-1.3547	mg/l	4/27/2023 11:46:28	
CCV2	S	Phenolics-	P	0.9328	mg/l	4/27/2023 11:46:29	
CCB2	S	Phenolics-	P	-0.0041	mg/l	4/27/2023 11:46:30	
O2521-05	S	Phenolics-	P	-0.7209	mg/l	4/27/2023 11:46:31	
O2521-07	S	Phenolics-	P	-0.5644	mg/l	4/27/2023 12:24:01	
O2521-07DUP	S	Phenolics-	P	-0.5577	mg/l	4/27/2023 12:24:04	
O2521-09MS	S	Phenolics-	P	0.4431	mg/l	4/27/2023 12:24:05	
O2521-10MSD	S	Phenolics-	P	0.4112	mg/l	4/27/2023 12:24:06	
O2521-14	S	Phenolics-	P	-0.0019	mg/l	4/27/2023 12:24:09	
CCV3	S	Phenolics-	P	0.9449	mg/l	4/27/2023 12:24:10	
CCB3	S	Phenolics-	P	-0.003	mg/l	4/27/2023 12:24:11	

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

4/27/2023 10:53


Test Phenolics-

Accepted 4/27/2023 10:53

Factor 9.322
Bias 0.008

Coeff. of det. 0.999981

Errors



	Calibrator	Response	Calc. con.	Conc.
1	0.00PPM	0.008	0.0033	0.0000
2	Phenol-2	0.013	0.0498	0.0500
3	Phenol-2	0.018	0.0975	0.1000
4	Phenol-2	0.034	0.2464	0.2500
5	Phenol-2	0.061	0.4998	0.5000
6	Phenol-2	0.115	1.0051	1.0000
7	Phenol-2	0.222	1.9981	2.0000

le
Errors

-
-0.4
-2.5
-1.4
0.0
0.5
-0.1

12
4/27/23