



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB125189

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/27/2023	14:30
2	CAL2	1	Water	NA	NA	20.7	7.00	04/27/2023	14:31
3	CAL3	1	Water	NA	NA	20.8	10.03	04/27/2023	14:33
4	ICV	1	Water	NA	NA	20.8	6.99	04/27/2023	14:35
5	CCV1	1	Water	NA	NA	20.8	2.01	04/27/2023	14:40
6	O2521-01	1	Water	NA	NA	20.8	6.85	04/27/2023	14:45
7	O2521-01DUP	1	Water	NA	NA	20.9	6.86	04/27/2023	14:46
8	O2521-03	1	Water	NA	NA	20.1	6.71	04/27/2023	14:49
9	O2521-05	1	Water	NA	NA	20.9	6.72	04/27/2023	15:00
10	O2521-07	1	Water	NA	NA	20.9	7.02	04/27/2023	15:05
11	O2521-14	1	Water	NA	NA	20.6	5.48	04/27/2023	15:10
12	CCV2	1	Water	NA	NA	20.7	12.02	04/27/2023	15:15