



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125194

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	04/27/2023	08:40
2	O2497-04		1	14	100.00	>212.0	>212.0	04/27/2023	09:10
3	O2497-04DUP		1	14	100.00	>212.0	>212.0	04/27/2023	09:40
4	O2497-05		1	13	100.00	>212.0	>212.0	04/27/2023	10:10
5	O2497-06		1	12	100.00	>212.0	>212.0	04/27/2023	10:40
6	O2508-03		1	14	100.00	>212.0	>212.0	04/27/2023	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66125194

WorkList Name : FP-4-

WorkList ID : 169436

Department : Wet-Chemistry

Date : 04-26-2023 16:29:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/03/2023	Water	O2497-04	Flash Point	Cool 4 deg C	PSEG03	M11	34279	04/26/2023	1010B
05/03/2023	Water	O2497-05	Flash Point	Cool 4 deg C	PSEG03	M11	34558	04/26/2023	1010B
05/03/2023	Water	O2497-06	Flash Point	Cool 4 deg C	PSEG03	M11	34564	04/26/2023	1010B
05/03/2023	Water	O2508-03	Flash Point	Cool 4 deg C	PSEG03	M11	34333	04/26/2023	1010B

Date/Time 04/27/2023 08:30
 Raw Sample Received by: Roy
 Raw Sample Relinquished by: JN(CDC)

Date/Time 04/27/2023 11:50
 Raw Sample Received by: JN(CDC)
 Raw Sample Relinquished by: Roy Cury