

LB125195


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

52620



A= 0.0000e+000

B= 1.9303e-004 slope

C= -5.0363e-003 y-intercept

Rho= 0.9990736

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.022	0.022	140	0.000	140					=
0.05	0.050					302	0				1
0.20	0.200	0.201	0.001	1067	0.0 %	1067					0
2.50	2.500	2.496	-0.004	12956	0.0 %	12956					3
5.00	5.000	5.152	0.152	26718	0.0 %	26718					4
7.50	7.500	7.177	-0.323	37206	0.0 %	37206					2
10.0	10.000	10.152	0.152	52620	0.0 %	52620					

LB125195

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	140	-		0 SFAM01.1	PPB	4/27/2023 15:59	Std
	0.2 S01	1067	-		0.2 SFAM01.1	PPB	4/27/2023 16:02	Std
	2.5 S02	12956	-		2.5 SFAM01.1	PPB	4/27/2023 16:04	Std
	5 S03	26718	-		5 SFAM01.1	PPB	4/27/2023 16:06	Std
	7.5 S04	37206	-		7.5 SFAM01.1	PPB	4/27/2023 16:08	Std
	10 S05	52620	-		10 SFAM01.1	PPB	4/27/2023 16:11	Std
ICV051	ICV051	21979	4.2375 -		SFAM01.1	PPB	4/27/2023 16:13	SMPL
ICB051	ICB051	-4	-0.0058 -		SFAM01.1	PPB	4/27/2023 16:16	SMPL
CCV075	CCV075	26590	5.1276 -		SFAM01.1	PPB	4/27/2023 16:18	SMPL
CCB075	CCB075	-24	-0.0097 -		SFAM01.1	PPB	4/27/2023 16:20	SMPL
PB152455BL	PBS	95	0.0133 -		SFAM01.1	PPB	4/27/2023 16:22	SMPL
O2517-01	BH4D8	2664	0.5092 -		SFAM01.1	PPB	4/27/2023 16:25	SMPL
O2517-02	BH4D8D	2281	0.4353 -		SFAM01.1	PPB	4/27/2023 16:27	SMPL
O2517-03	BH4D8S	16935	3.2639 -		SFAM01.1	PPB	4/27/2023 16:29	SMPL
O2517-04	BH4D9	2160	0.4119 -		SFAM01.1	PPB	4/27/2023 16:31	SMPL
O2517-05	BH4E0	2044	0.3895 -		SFAM01.1	PPB	4/27/2023 16:34	SMPL
O2517-06	BH4E1	2105	0.4013 -		SFAM01.1	PPB	4/27/2023 16:36	SMPL
CCV076	CCV076	26358	5.0828 -		SFAM01.1	PPB	4/27/2023 16:38	SMPL
CCB076	CCB076	14	-0.0023 -		SFAM01.1	PPB	4/27/2023 16:41	SMPL