

LB125203


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

51123



A= 0.0000e+000

B= 1.9586e-004 slope

C= -3.8734e-002 y -intercept

Rho= 0.9999640

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.021	-0.021	92	0.000	92					
0.05	0.050					302	0				
0.20	0.200	0.233	0.033	1386	0.0 %	1386					
2.50	2.500	2.483	-0.017	12875	0.0 %	12875					
5.00	5.000	4.977	-0.023	25610	0.0 %	25610					
7.50	7.500	7.554	0.054	38766	0.0 %	38766					
10.0	10.000	9.974	-0.026	51123	0.0 %	51123					

100-0-0

LB125203
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	92	-		0 SFAM01.1	PPB	4/28/2023 11:47	Std
	0.2 S01	1386	-		0.2 SFAM01.1	PPB	4/28/2023 11:49	Std
	2.5 S02	12875	-		2.5 SFAM01.1	PPB	4/28/2023 11:52	Std
	5 S03	25610	-		5 SFAM01.1	PPB	4/28/2023 11:54	Std
	7.5 S04	38766	-		7.5 SFAM01.1	PPB	4/28/2023 11:56	Std
	10 S05	51123	-		10 SFAM01.1	PPB	4/28/2023 11:58	Std
ICV052	ICV052	21882	4.247 -		SFAM01.1	PPB	4/28/2023 12:01	SMPL
ICB052	ICB052	-48	-0.0481 -		SFAM01.1	PPB	4/28/2023 12:03	SMPL
CCV077	CCV077	25531	4.9617 -		SFAM01.1	PPB	4/28/2023 12:06	SMPL
CCB077	CCB077	-54	-0.0493 -		SFAM01.1	PPB	4/28/2023 12:08	SMPL
PB152475BL	PBW475	36	-0.0317 -		SFAM01.1	PPB	4/28/2023 12:10	SMPL
O2376-01	MCB9D2	70	-0.025 -		SFAM01.1	PPB	4/28/2023 12:12	SMPL
O2376-02	MCB9D4	77	-0.0237 -		SFAM01.1	PPB	4/28/2023 12:15	SMPL
O2376-03	MCB9D4D	73	-0.0244 -		SFAM01.1	PPB	4/28/2023 12:17	SMPL
O2376-04	MCB9D4S	4383	0.8197 -		SFAM01.1	PPB	4/28/2023 12:19	SMPL
CCV078	CCV078	25539	4.9633 -		SFAM01.1	PPB	4/28/2023 12:22	SMPL
CCB078	CCB078	-87	-0.0558 -		SFAM01.1	PPB	4/28/2023 12:24	SMPL