


7470A

LB125204
INSTRUMENT ID: CV1

Linear

 μ Abs.:

4464E



A= 0.0000e+000

B= 2.2909e-004

C= -2.6594e-002

Rho= 0.9972597

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	-0.008	-0.008	82	0.000	82	0				
0.2	0.200	0.227	0.027	1109	0.0 %	1109					
2.5	2.500	2.498	-0.002	11021	0.0 %	11021					
5.0	5.000	5.327	0.327	23369	0.0 %	23369					
7.5	7.500	6.953	-0.547	30468	0.0 %	30468					
10.0	10.000	10.202	0.202	44646	0.0 %	44646					

0/00
14
0
7
7
2

Method: 7470A

Operator: Admin

LB125204 INSTRUMENT ID : CV1

Date of Analysis: 28 Apr 2023 12:38:24

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SP	82	-	0.0000	7470A	PPB	28 Apr 2023 12:49:48	S	Std
0.2 - 1	SP-2	1109	-	0.2000	7470A	PPB	28 Apr 2023 12:52:06	S	Std
2.5 - 1	SP-5	11021	-	2.5000	7470A	PPB	28 Apr 2023 12:54:24	S	Std
5.0 - 1	SP	23369	-	5.0000	7470A	PPB	28 Apr 2023 12:56:41	S	Std
7.5 - 1	SP-1	30468	-	7.5000	7470A	PPB	28 Apr 2023 12:58:57	S	Std
10.0 - 1	SP-1	44646	-	10.0000	7470A	PPB	28 Apr 2023 13:01:14	S	Std
ICV53 - 1	ICV53	19279	4.3901	-	7470A	PPB	28 Apr 2023 13:11:45	U	SMPL
ICB53 - 1	ICB53	-132	-0.0568	-	7470A	PPB	28 Apr 2023 13:14:00	U	SMPL
CCV79 - 1	CCV79	22491	5.1260	-	7470A	PPB	28 Apr 2023 13:16:17	U	SMPL
CCB79 - 1	CCB79	-127	-0.0557	-	7470A	PPB	28 Apr 2023 13:18:32	U	SMPL
CRA - 1	CRA	1035	0.2105	-	7470A	PPB	28 Apr 2023 13:20:49	U	SMPL
HighStd - 1	HighStd	43901	10.0309	-	7470A	PPB	28 Apr 2023 13:23:04	U	SMPL
ChkStd - 1	ChkStd	26576	6.0618	-	7470A	PPB	28 Apr 2023 13:25:20	U	SMPL
PB152487BL - 1	PBW	-286	-0.0921	-	7470A	PPB	28 Apr 2023 13:27:38	U	SMPL
PB152487BS - 1	LCSW	16243	3.6946	-	7470A	PPB	28 Apr 2023 13:29:57	U	SMPL
O2495-01 - 1	PZ-01	1810	0.3881	-	7470A	PPB	28 Apr 2023 13:32:13	U	SMPL
O2495-03 - 1	PZ-02	3344	0.7395	-	7470A	PPB	28 Apr 2023 13:34:30	U	SMPL
O2495-05 - 1	GMW-11	2585	0.5656	-	7470A	PPB	28 Apr 2023 13:36:49	U	SMPL
O2495-07 - 1	GMW-X	2233	0.4850	-	7470A	PPB	28 Apr 2023 13:39:07	U	SMPL
O2495-09 - 1	GMW-5	82	-0.0078	-	7470A	PPB	28 Apr 2023 13:41:25	U	SMPL
CCV80 - 1	CCV80	21839	4.9766	-	7470A	PPB	28 Apr 2023 13:43:41	U	SMPL
CCB80 - 1	CCB80	-65	-0.0415	-	7470A	PPB	28 Apr 2023 13:45:57	U	SMPL
O2496-01 - 1	TOTE-4-25-2023	154	0.0087	-	7470A	PPB	28 Apr 2023 13:48:12	U	SMPL
O2497-05 - 1	34558	222	0.0243	-	7470A	PPB	28 Apr 2023 13:50:28	U	SMPL
O2521-01 - 1	GMW-1	1052	0.2144	-	7470A	PPB	28 Apr 2023 13:52:43	U	SMPL
O2521-03 - 1	PZ-15	3933	0.8744	-	7470A	PPB	28 Apr 2023 13:54:59	U	SMPL
O2521-05 - 1	GMW-10	266	0.0343	-	7470A	PPB	28 Apr 2023 13:57:17	U	SMPL
O2521-07 - 1	GMW-7	478	0.0829	-	7470A	PPB	28 Apr 2023 13:59:35	U	SMPL
O2521-07DUP - 1	GMW-7DUP	507	0.0896	-	7470A	PPB	28 Apr 2023 14:01:52	U	SMPL
O2521-09 - 1	O2521-07MS	20269	4.6169	-	7470A	PPB	28 Apr 2023 14:07:16	U	SMPL
O2521-10 - 1	O2521-07MSD	19134	4.3569	-	7470A	PPB	28 Apr 2023 14:09:32	U	SMPL
O2521-14 - 1	FB	-81	-0.0452	-	7470A	PPB	28 Apr 2023 14:11:49	U	SMPL
CCV81 - 1	CCV81	21678	4.9397	-	7470A	PPB	28 Apr 2023 14:14:05	U	SMPL
CCB81 - 1	CCB81	-128	-0.0559	-	7470A	PPB	28 Apr 2023 14:16:21	U	SMPL
O2485-02 - 1	33994	3738	0.8298	-	7470A	PPB	28 Apr 2023 14:18:38	U	SMPL
O2521-07LX5 - 1		-202	-0.0729	-	7470A	PPB	28 Apr 2023 14:20:54	U	SMPL
O2521-07A - 1		19259	4.3855	-	7470A	PPB	28 Apr 2023 14:28:33	U	SMPL
CCV82 - 1	CCV82	22062	5.0277	-	7470A	PPB	28 Apr 2023 14:30:48	U	SMPL
CCB82 - 1	CCB82	-115	-0.0529	-	7470A	PPB	28 Apr 2023 14:33:05	U	SMPL