

LB 125213


SFAM01.1

INSTRUMENT ID: CVL

Linear

 μ Abs.:

51503



A= 0.0000e+000

B= 1.9768e-004 slope

C= 1.3680e-002 y-intercept

Rho= 0.9989573

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/d
0.00	0.000	0.039	0.039	130	0.000	130					
0.05	0.050					302	0				
0.20	0.200	0.215	0.015	1016	0.0 %	1016					
2.50	2.500	2.503	0.003	12593	0.0 %	12593					
5.00	5.000	5.096	0.096	25711	0.0 %	25711					
7.50	7.500	7.152	-0.348	36109	0.0 %	36109					
10.0	10.000	10.195	0.195	51503	0.0 %	51503					

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	130	-		0 SFAM01.1	PPB	4/28/2023 15:03	Std
	0.2 S01	1016	-		0.2 SFAM01.1	PPB	4/28/2023 15:05	Std
	2.5 S02	12593	-		2.5 SFAM01.1	PPB	4/28/2023 15:10	Std
	5 S03	25711	-		5 SFAM01.1	PPB	4/28/2023 15:12	Std
	7.5 S04	36109	-		7.5 SFAM01.1	PPB	4/28/2023 15:15	Std
	10 S05	51503	-		10 SFAM01.1	PPB	4/28/2023 15:17	Std
ICV054	ICV054	22567	4.3812 -		SFAM01.1	PPB	4/28/2023 15:20	SMPL
ICB054	ICB054	0	-0.0387 -		SFAM01.1	PPB	4/28/2023 15:22	SMPL
CCV083	CCV083	26110	5.0751 -		SFAM01.1	PPB	4/28/2023 15:24	SMPL
CCB083	CCB083	5	-0.0378 -		SFAM01.1	PPB	4/28/2023 15:26	SMPL
PB152476BL	PBS476	69	-0.0252 -		SFAM01.1	PPB	4/28/2023 15:29	SMPL
O2377-01	MCB9D5	6254	1.1862 -		SFAM01.1	PPB	4/28/2023 15:31	SMPL
O2377-02	MCB9D6	4496	0.8418 -		SFAM01.1	PPB	4/28/2023 15:33	SMPL
O2377-03	MCB9D6D	4757	0.893 -		SFAM01.1	PPB	4/28/2023 15:35	SMPL
O2377-04	MCB9D6S	18241	3.5339 -		SFAM01.1	PPB	4/28/2023 15:38	SMPL
PB152477BL	PBS477	-51	-0.0487 -		SFAM01.1	PPB	4/28/2023 15:40	SMPL
O2480-01	MH2849	957	0.1487 -		SFAM01.1	PPB	4/28/2023 15:42	SMPL
O2480-02	MH2849D	961	0.1495 -		SFAM01.1	PPB	4/28/2023 15:44	SMPL
O2480-03	MH2849S	14437	2.7889 -		SFAM01.1	PPB	4/28/2023 15:47	SMPL
O2480-04	MH2850	19686	3.8169 -		SFAM01.1	PPB	4/28/2023 15:49	SMPL
O2480-05	MH2851	933	0.144 -		SFAM01.1	PPB	4/28/2023 15:51	SMPL
O2480-06	MH2852	1101	0.1769 -		SFAM01.1	PPB	4/28/2023 15:54	SMPL
O2480-07	MH2853	845	0.1268 -		SFAM01.1	PPB	4/28/2023 15:56	SMPL
O2480-08	MH2854	7126	1.3569 -		SFAM01.1	PPB	4/28/2023 15:58	SMPL
O2480-09	MH2855	3676	0.6812 -		SFAM01.1	PPB	4/28/2023 16:00	SMPL
O2480-10	MH2856	13140	2.5348 -		SFAM01.1	PPB	4/28/2023 16:03	SMPL
O2480-11	MH2857	2143	0.381 -		SFAM01.1	PPB	4/28/2023 16:05	SMPL
O2480-12	MH2858	2078	0.3683 -		SFAM01.1	PPB	4/28/2023 16:07	SMPL
O2480-13	MH2859	2073	0.3673 -		SFAM01.1	PPB	4/28/2023 16:09	SMPL
O2480-14	MH2860	806	0.1191 -		SFAM01.1	PPB	4/28/2023 16:12	SMPL
CCV084	CCV084	26496	5.1507 -		SFAM01.1	PPB	4/28/2023 16:14	SMPL
CCB084	CCB084	2	-0.0383 -		SFAM01.1	PPB	4/28/2023 16:16	SMPL
O2480-15	MH2861	3189	0.5859 -		SFAM01.1	PPB	4/28/2023 16:19	SMPL
O2480-16	MH2862	5197	0.9791 -		SFAM01.1	PPB	4/28/2023 16:21	SMPL
O2480-17	MH2863	1485	0.2521 -		SFAM01.1	PPB	4/28/2023 16:23	SMPL
O2480-18	MH2864	1099	0.1765 -		SFAM01.1	PPB	4/28/2023 16:25	SMPL
O2480-19	MH2865	424	0.0443 -		SFAM01.1	PPB	4/28/2023 16:28	SMPL
O2480-20	MH2866	388	0.0373 -		SFAM01.1	PPB	4/28/2023 16:30	SMPL
O2480-21	MH2867	330	0.0259 -		SFAM01.1	PPB	4/28/2023 16:32	SMPL
O2480-22	MH2868	7898	1.5081 -		SFAM01.1	PPB	4/28/2023 16:34	SMPL
CCV085	CCV085	26311	5.1145 -		SFAM01.1	PPB	4/28/2023 16:37	SMPL
CCB085	CCB085	-23	-0.0432 -		SFAM01.1	PPB	4/28/2023 16:39	SMPL