

LB125227


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

44454



A= 0.0000e+000

B= 2.2357e-004

C= 1.9327e-002

Rho= 0.9997436

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	0.032	0.032	56	0.000	56	0				18
0.2	0.200	0.236	0.036	969	0.0 %	969					-7
2.5	2.500	2.337	-0.163	10368	0.0 %	10368					2
5.0	5.000	5.099	0.099	22723	0.0 %	22723					1
7.5	7.500	7.538	0.038	33630	0.0 %	33630					0
10.0	10.000	9.958	-0.042	44454	0.0 %	44454					

LB125227 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 May 2023 13:48:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	56	-	0.000	7470A	PPB	01 May 2023 13:51:13	S	Std
0.2 - 1	50.2	969	-	0.200	7470A	PPB	01 May 2023 13:53:30	S	Std
2.5 - 1	52.5	10368	-	2.500	7470A	PPB	01 May 2023 13:55:47	S	Std
5.0 - 1	55	22723	-	5.000	7470A	PPB	01 May 2023 13:58:04	S	Std
7.5 - 1	57.5	33630	-	7.500	7470A	PPB	01 May 2023 14:03:03	S	Std
10.0 - 1	510	44454	-	10.000	7470A	PPB	01 May 2023 14:05:19	S	Std
ICV55 - 1	ICV55	18597	4.1770	-	7470A	PPB	01 May 2023 14:10:30	U	SMPL
ICB55 - 1	ICB55	-197	-0.0247	-	7470A	PPB	01 May 2023 14:12:45	U	SMPL
CCV86 - 1	CCV86	23304	5.2293	-	7470A	PPB	01 May 2023 14:15:02	U	SMPL
CCB86 - 1	CCB86	-165	-0.0176	-	7470A	PPB	01 May 2023 14:17:17	U	SMPL
CRA - 1	CRA	980	0.2384	-	7470A	PPB	01 May 2023 14:19:35	U	SMPL
HighStd - 1	HighStd	44914	10.0605	-	7470A	PPB	01 May 2023 14:21:50	U	SMPL
ChkStd - 1	ChkStd	32062	7.1873	-	7470A	PPB	01 May 2023 14:24:05	U	SMPL
PB152503BL - 1	PBW	-88	-0.0003	-	7470A	PPB	01 May 2023 14:26:24	U	SMPL
PB152503BS - 1	LCSW	18495	4.1542	-	7470A	PPB	01 May 2023 14:28:41	U	SMPL
O2416-11 - 1	NATIONAL-GRID-COMP	-64	0.0050	-	7470A	PPB	01 May 2023 14:30:58	U	SMPL
O2476-04 - 1	M-041923RB04	245	0.0741	-	7470A	PPB	01 May 2023 14:33:15	U	SMPL
O2501-01 - 1	RPV-ETGI-352	105	0.0428	-	7470A	PPB	01 May 2023 14:35:33	U	SMPL
O2502-02 - 1	BU-511-COMP-01	250	0.0752	-	7470A	PPB	01 May 2023 14:37:50	U	SMPL
O2502-09 - 1	BU-511-COMP-02	99	0.0415	-	7470A	PPB	01 May 2023 14:40:06	U	SMPL
CCV87 - 1	CCV87	23939	5.3713	-	7470A	PPB	01 May 2023 14:42:21	U	SMPL
CCB87 - 1	CCB87	-154	-0.0151	-	7470A	PPB	01 May 2023 14:44:37	U	SMPL
O2503-02 - 1	TR-305A-COMP-01	90	0.0394	-	7470A	PPB	01 May 2023 14:46:54	U	SMPL
O2503-09 - 1	TR-305A-COMP-02	91	0.0397	-	7470A	PPB	01 May 2023 14:49:09	U	SMPL
O2503-16 - 1	TR-305A-COMP-03	180	0.0596	-	7470A	PPB	01 May 2023 14:51:25	U	SMPL
O2504-02 - 1	BU-512-COMP-01	187	0.0611	-	7470A	PPB	01 May 2023 14:53:41	U	SMPL
O2504-09 - 1	BU-512-COMP-02	143	0.0513	-	7470A	PPB	01 May 2023 14:55:57	U	SMPL
O2504-16 - 1	BU-512-COMP-03	173	0.0580	-	7470A	PPB	01 May 2023 14:58:14	U	SMPL
O2504-16DUP - 1	BU-512-COMP-03DUP	236	0.0721	-	7470A	PPB	01 May 2023 15:00:30	U	SMPL
O2504-16MS - 1	BU-512-COMP-03MS	16340	3.6724	-	7470A	PPB	01 May 2023 15:02:48	U	SMPL
O2504-16MSD - 1	BU-512-COMP-03MSD	16720	3.7573	-	7470A	PPB	01 May 2023 15:05:05	U	SMPL
CCV88 - 1	CCV88	21762	4.8846	-	7470A	PPB	01 May 2023 15:12:40	U	SMPL
CCB88 - 1	CCB88	-168	-0.0182	-	7470A	PPB	01 May 2023 15:14:55	U	SMPL
O2504-16LX5 - 1		18	0.0234	-	7470A	PPB	01 May 2023 15:17:12	U	SMPL
O2504-16A - 1		17869	4.0142	-	7470A	PPB	01 May 2023 15:19:28	U	SMPL
PB152433TB - 1	PB152433TB	-9	0.0173	-	7470A	PPB	01 May 2023 15:21:44	U	SMPL
CCV89 - 1	CCV89	23167	5.1987	-	7470A	PPB	01 May 2023 15:24:01	U	SMPL
CCB89 - 1	CCB89	-145	-0.0131	-	7470A	PPB	01 May 2023 15:26:16	U	SMPL