



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB125228
Analysis Date: 05/01/2023
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 05/01/2023
Extraction IN Time: 14:10
Extraction OUT Time: 14:50
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB125228BL	LB125228BL	WATER	1.3	1000	100	3.0503	3.0503	3.02	3.0504	3.0504	0.0001	0.1
2	LB125228BS	LB125228BS	WATER	1.3	1000	100	3.1487	3.1487	3.01	3.1656	3.1656	0.0169	16.9
3	LB125228BSD	LB125228BSD	WATER	1.3	1000	100	3.1960	3.1960	3.03	3.2130	3.2130	0.0170	17
4	O2498-01	GRAB	WATER	1.6	1000	100	3.0082	3.0082	3.04	3.0669	3.0669	0.0587	58.7
5	O2529-02	WWWC	WATER	1.6	1000	100	3.1105	3.1105	3.06	3.1192	3.1192	0.0087	8.7



QC Batch# LB125228

Test: TPH

Analysis Date: 05/01/2023

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W3010
pH Paper 0-14	M4909
Sodium Sulfate	EP2228
1:1 HCL	WP101334
Silica Gel	W2839
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP101332
LCSWD	5.00 ML	WP101333
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0019 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 16:01

1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 15:15

Bal Check Time: 14:30 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 16:30

Out Time1: 16:00

After Analysis

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 17:31

1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time2: 17:00

Bal Check Time: 18:05 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 18:00

Out Time2: 17:30

WORKLIST(Hardcopy Internal Chain)

125228

WorkList Name : tph o2498

WorkList ID : 169558

Department : Wet-Chemistry

Date : 05-01-2023 13:35:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/09/2023	Water	O2498-01	TPH	Conc H2SO4 to	ARAM01	N11	GRAB	04/26/2023	1664A
05/04/2023	Water	O2529-02	TPH	Conc H2SO4 to	DVIR01	N11	WWWC	04/26/2023	1664A

Date/Time 05/01/23 14:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 05/01/23 16:35

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]