



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB125229

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	05/01/2023	13:42
2	CAL2	1	Water	NA	NA	20.7	7.00	05/01/2023	13:43
3	CAL3	1	Water	NA	NA	20.7	10.03	05/01/2023	13:44
4	ICV	1	Water	NA	NA	20.8	6.99	05/01/2023	13:50
5	CCV1	1	Water	NA	NA	20.7	2.01	05/01/2023	14:00
6	O2552-01	1	Water	NA	NA	24.1	11.77	05/01/2023	14:10
7	O2552-03	1	Water	NA	NA	24.4	11.82	05/01/2023	14:15
8	O2556-03	1	Water	NA	NA	20.9	6.82	05/01/2023	14:25
9	O2556-03DUP	1	Water	NA	NA	20.8	6.83	05/01/2023	14:26
10	CCV2	1	Water	NA	NA	20.7	12.03	05/01/2023	14:30

NR 125229

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 02556 WorkList ID : 169559 Department : Wet-Chemistry Date : 05-01-2023 13:36:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/28/2023	Water	O2552-01	pH	Cool 4 deg C	PSEG03	M11	SAYR-1120	04/28/2023	9040C
04/28/2023	Water	O2552-03	pH	Cool 4 deg C	PSEG03	M11	SAYR-1121	04/28/2023	9040C
04/28/2023	Water	O2556-03	pH	Cool 4 deg C	PSEG03	K11	MW-1-4-WATER	04/28/2023	9040C

Date/Time 05/01/23 13:40
 Raw Sample Received by: 70(WOC)
 Raw Sample Relinquished by: [Signature]

Date/Time 05/01/23 16:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: 70(COC)