

LB125235


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51287



A= 0.0000e+000

B= 1.9871e-004 *51287*C= 2.9953e-002 *1=100 for rep*

Rho= 0.9987179

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.043	0.043	66	0.000	66					=
0.05	0.050					302	0				16
0.20	0.200	0.232	0.032	1018	0.0 %	1018					0
2.50	2.500	2.493	-0.007	12394	0.0 %	12394					2
5.00	5.000	5.096	0.096	25495	0.0 %	25495					-5
7.50	7.500	7.115	-0.385	35655	0.0 %	35655					2
10.0	10.000	10.221	0.221	51287	0.0 %	51287					

LB125235

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	66	-		0 SFAM01.1	PPB	5/1/2023 16:08	Std
	0.2 S01	1018	-		0.2 SFAM01.1	PPB	5/1/2023 16:10	Std
	2.5 S02	12394	-		2.5 SFAM01.1	PPB	5/1/2023 16:13	Std
	5 S03	25495	-		5 SFAM01.1	PPB	5/1/2023 16:15	Std
	7.5 S04	35655	-		7.5 SFAM01.1	PPB	5/1/2023 16:17	Std
	10 S05	51287	-		10 SFAM01.1	PPB	5/1/2023 16:20	Std
ICV056	ICV056	20618	4.1269 -		SFAM01.1	PPB	5/1/2023 16:25	SMPL
ICB056	ICB056	-68	0.0164 -		SFAM01.1	PPB	5/1/2023 16:27	SMPL
CCV089	CCV89	25563	5.1095 -		SFAM01.1	PPB	5/1/2023 16:29	SMPL
CCB089	CCB89	-88	0.0125 -		SFAM01.1	PPB	5/1/2023 16:32	SMPL
PB152476BL	PBS476	13	0.0325 -		SFAM01.1	PPB	5/1/2023 16:34	SMPL
O2377-01	MCB9D5	5750	1.1725 -		SFAM01.1	PPB	5/1/2023 16:36	SMPL
O2377-02	MCB9D6	4251	0.8747 -		SFAM01.1	PPB	5/1/2023 16:38	SMPL
O2377-03	MCB9D6D	4565	0.9371 -		SFAM01.1	PPB	5/1/2023 16:43	SMPL
O2377-04	MCB9D6S	17208	3.4493 -		SFAM01.1	PPB	5/1/2023 16:45	SMPL
PB152477BL	PBS477	-88	0.0125 -		SFAM01.1	PPB	5/1/2023 16:47	SMPL
O2480-01	MH2849	951	0.2189 -		SFAM01.1	PPB	5/1/2023 16:50	SMPL
O2480-02	MH2849D	883	0.2054 -		SFAM01.1	PPB	5/1/2023 16:52	SMPL
O2480-03	MH2849S	13911	2.7942 -		SFAM01.1	PPB	5/1/2023 16:54	SMPL
O2480-04	MH2850	19213	3.8477 -		SFAM01.1	PPB	5/1/2023 16:57	SMPL
O2480-05	MH2851	890	0.2068 -		SFAM01.1	PPB	5/1/2023 16:59	SMPL
O2480-06	MH2852	1041	0.2368 -		SFAM01.1	PPB	5/1/2023 17:01	SMPL
O2480-07	MH2853	702	0.1694 -		SFAM01.1	PPB	5/1/2023 17:03	SMPL
O2480-08	MH2854	6887	1.3985 -		SFAM01.1	PPB	5/1/2023 17:06	SMPL
O2480-09	MH2855	3163	0.6585 -		SFAM01.1	PPB	5/1/2023 17:08	SMPL
O2480-10	MH2856	12223	2.4588 -		SFAM01.1	PPB	5/1/2023 17:10	SMPL
O2480-11	MH2857	2018	0.4309 -		SFAM01.1	PPB	5/1/2023 17:13	SMPL
O2480-12	MH2858	1637	0.3552 -		SFAM01.1	PPB	5/1/2023 17:15	SMPL
O2480-13	MH2859	1789	0.3854 -		SFAM01.1	PPB	5/1/2023 17:17	SMPL
O2480-14	MH2860	765	0.182 -		SFAM01.1	PPB	5/1/2023 17:19	SMPL
CCV090	CCV090	25803	5.1572 -		SFAM01.1	PPB	5/1/2023 17:22	SMPL
CCB090	CCB090	-83	0.0135 -		SFAM01.1	PPB	5/1/2023 17:24	SMPL
O2480-15	MH2861	3162	0.6583 -		SFAM01.1	PPB	5/1/2023 17:26	SMPL
O2480-16	MH2862	5073	1.038 -		SFAM01.1	PPB	5/1/2023 17:28	SMPL
O2480-17	MH2863	1471	0.3223 -		SFAM01.1	PPB	5/1/2023 17:31	SMPL
O2480-18	MH2864	1084	0.2454 -		SFAM01.1	PPB	5/1/2023 17:33	SMPL
O2480-19	MH2865	405	0.1104 -		SFAM01.1	PPB	5/1/2023 17:35	SMPL
O2480-20	MH2866	327	0.0949 -		SFAM01.1	PPB	5/1/2023 17:38	SMPL
O2480-21	MH2867	270	0.0836 -		SFAM01.1	PPB	5/1/2023 17:40	SMPL
O2480-22	MH2868	7491	1.5185 -		SFAM01.1	PPB	5/1/2023 17:42	SMPL
CCV091	CCV091	23627	4.7248 -		SFAM01.1	PPB	5/1/2023 17:49	SMPL
CCB091	CCB091	-69	0.0162 -		SFAM01.1	PPB	5/1/2023 17:51	SMPL