



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125237

Ambient Barometric Pressure (mmHg): 745.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	83.3	05/01/2023	13:45
2	O2497-02		1	14	100.00	>212.0	>212.0	05/01/2023	14:15
3	O2508-01		1	12	100.00	>212.0	>212.0	05/01/2023	14:45
4	O2519-04		1	12	57.00	134.6	135.5	05/01/2023	15:15
5	O2519-04DUP		1	12	57.00	134.6	135.5	05/01/2023	15:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

26125237

WorkList Name : FP 4-W

WorkList ID : 169437

Department : Wet-Chemistry

Date : 04-26-2023 16:30:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/03/2023	Solid	02497-02	Flash Point	Cool 4 deg C	PSEG03	M11	32498	04/26/2023	1010B
05/03/2023	Solid	02508-01	Flash Point	Cool 4 deg C	PSEG03	M11	32514	04/26/2023	1010B
05/03/2023	Solid	02519-04	Flash Point	Cool 4 deg C	PSEG03	M11	18827	04/26/2023	1010B

Date/Time 05/01/2023 13:10
 Raw Sample Received by: RFA Curo
 Raw Sample Relinquished by: JHCOCI

Date/Time 05/01/2023 16:26
 Raw Sample Received by: JHCOCI
 Raw Sample Relinquished by: RFA Curo