



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/3/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 15:50
In Date: 05/02/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:34
Out Date: 05/03/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125246

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2578-01	EYXL77	1	1.18	8.60	9.78	7.94	78.6	
O2578-02	EYXL78	2	1.12	8.65	9.77	7.87	78.0	
O2578-03	EYXL79	3	1.17	8.54	9.71	7.88	78.6	
O2578-04	EYXL80	4	1.19	8.78	9.97	8.11	78.8	
O2578-05	EYXL81	5	1.14	8.59	9.73	7.9	78.7	
O2578-06	EYXL81D	6	1.14	8.59	9.73	7.9	78.7	
O2578-07	EYXL81S	7	1.14	8.59	9.73	7.9	78.7	
O2578-08	EYXL82	8	1.19	8.58	9.77	7.82	77.3	
O2578-09	EYXL83	9	1.19	8.60	9.79	7.37	71.9	
O2578-10	EYXL84	10	1.1	8.61	9.71	7.66	76.2	
O2578-11	EYXL85	11	1.1	8.44	9.54	7.69	78.1	
O2578-12	EYXL86	12	1.19	8.78	9.97	8.12	78.9	
O2578-13	EYXL87	13	1.19	8.65	9.84	7.9	77.6	
O2578-14	EYXL88	14	1.16	8.78	9.94	7.52	72.4	
O2578-15	EYXL89	15	1.15	8.62	9.77	7.6	74.8	
O2578-16	EYXL90	16	1.19	8.58	9.77	7.62	74.9	
O2578-17	EYXL91	17	1.13	8.58	9.71	7.65	76.0	
O2578-18	EYXL92	18	1.16	8.51	9.67	7.72	77.1	
O2578-19	EYXL93	19	1.17	8.81	9.98	8.69	85.4	
O2578-20	EYXL94	20	1.13	8.77	9.9	8.39	82.8	
O2578-21	EYXL95	21	1.11	8.77	9.88	8.17	80.5	
O2578-22	EYXL96	22	1.15	8.82	9.97	8.41	82.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$