

69125251


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

42514



A= 0.0000e+000

B= 2.3139e-004

C= 1.3786e-002

Rho= 0.9994091

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.033	0.033	84	0.000	84	0				-
0.2	0.200	0.182	-0.018	727	0.0 %	727					-9
2.5	2.500	2.463	-0.037	10586	0.0 %	10586					-1
5.0	5.000	4.916	-0.084	21186	0.0 %	21186					-2
7.5	7.500	7.754	0.254	33451	0.0 %	33451					3
10.0	10.000	9.851	-0.149	42514	0.0 %	42514					-1

LB125251 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 May 2023 15:26:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	84	-	0.0000	7470A	PPB	02 May 2023 15:51:57	S	Std
0.2 - 1	50.2	727	-	0.2000	7470A	PPB	02 May 2023 15:54:14	S	Std
2.5 - 1	52.5	10586	-	2.5000	7470A	PPB	02 May 2023 15:56:31	S	Std
5.0 - 1	55	21186	-	5.0000	7470A	PPB	02 May 2023 15:58:48	S	Std
7.5 - 1	59.5	33451	-	7.5000	7470A	PPB	02 May 2023 16:01:05	S	Std
10.0 - 1	510	42514	-	10.0000	7470A	PPB	02 May 2023 16:03:23	S	Std
ICV58 - 1	ICV58	16184	3.7587	-	7470A	PPB	02 May 2023 16:06:13	U	SMPL
ICB58 - 1	ICB58	-119	-0.0137	-	7470A	PPB	02 May 2023 16:08:28	U	SMPL
CCV98 - 1	CCV98	20742	4.8133	-	7470A	PPB	02 May 2023 16:10:45	U	SMPL
CCB98 - 1	CCB98	-115	-0.0128	-	7470A	PPB	02 May 2023 16:13:00	U	SMPL
CRA - 1	CRA	649	0.1640	-	7470A	PPB	02 May 2023 16:15:16	U	SMPL
HighStd - 1	HighStd	41284	9.5666	-	7470A	PPB	02 May 2023 16:17:32	U	SMPL
ChkStd - 1	ChkStd	29695	6.8850	-	7470A	PPB	02 May 2023 16:19:47	U	SMPL
PB152538BL - 1	PBW	-236	-0.0408	-	7470A	PPB	02 May 2023 16:22:06	U	SMPL
PB152538BS - 1	LCSW	15189	3.5284	-	7470A	PPB	02 May 2023 16:24:24	U	SMPL
O2495-02 - 1	PZ-01	-75	-0.0036	-	7470A	PPB	02 May 2023 16:26:40	U	SMPL
O2495-04 - 1	PZ-02	70	0.0300	-	7470A	PPB	02 May 2023 16:28:57	U	SMPL
O2495-06 - 1	GMW-11	49	0.0251	-	7470A	PPB	02 May 2023 16:31:15	U	SMPL
O2495-08 - 1	GMW-X	104	0.0379	-	7470A	PPB	02 May 2023 16:33:32	U	SMPL
O2495-10 - 1	GMW-5	77	0.0316	-	7470A	PPB	02 May 2023 16:35:47	U	SMPL
CCV99 - 1	CCV99	20594	4.7791	-	7470A	PPB	02 May 2023 16:38:02	U	SMPL
CCB99 - 1	CCB99	-135	-0.0175	-	7470A	PPB	02 May 2023 16:40:18	U	SMPL
O2521-02 - 1	GMW-1	46	0.0244	-	7470A	PPB	02 May 2023 16:42:35	U	SMPL
O2521-04 - 1	PZ-15	85	0.0335	-	7470A	PPB	02 May 2023 16:44:51	U	SMPL
O2521-06 - 1	GMW-10	42	0.0235	-	7470A	PPB	02 May 2023 16:47:06	U	SMPL
O2521-08 - 1	GMW-7	54	0.0263	-	7470A	PPB	02 May 2023 16:49:22	U	SMPL
O2521-08DUP - 1	GMW-7DUP	37	0.0223	-	7470A	PPB	02 May 2023 16:51:39	U	SMPL
O2521-11 - 1	O2521-08MS	15533	3.6080	-	7470A	PPB	02 May 2023 16:53:55	U	SMPL
O2521-12 - 1	O2521-08MSD	15056	3.4976	-	7470A	PPB	02 May 2023 16:56:12	U	SMPL
O2521-15 - 1	FB	-67	-0.0017	-	7470A	PPB	02 May 2023 16:58:30	U	SMPL
O2521-08LX5 - 1		93	0.0353	-	7470A	PPB	02 May 2023 17:00:48	U	SMPL
O2521-08A - 1		17536	4.0715	-	7470A	PPB	02 May 2023 17:03:04	U	SMPL
CCV01 - 1	CCV01	20305	4.7122	-	7470A	PPB	02 May 2023 17:05:20	U	SMPL
CCB01 - 1	CCB01	-141	-0.0188	-	7470A	PPB	02 May 2023 17:07:36	U	SMPL