



## Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB125253

| Reagent/Standard                        | Lot/Log # |
|-----------------------------------------|-----------|
| TURBIDITY LOD STD, 0.5NTU               | WP102259  |
| Turbidity Calibration std, 0NTU         | WP102252  |
| Turbidity Calibration std, 20NTU        | WP102254  |
| Turbidity Calibration std, 40NTU        | WP102253  |
| Turbidity Calibration std, 1NTU         | WP102257  |
| Turbidity Calibration std, 5NTU         | WP102256  |
| Turbidity LOQ std, 1.0NTU               | WP102260  |
| Turbidity Calibration - CCV std, 10 NTU | WP102255  |
| 10 NTU Standard 500 ml                  | W2971     |

Intercept: 0.026

Slope: 0.9369

Regression: 0.9999

| Seq | Lab ID | True Value (NTU) | Dilution | Reading | Result (mg/l) | %D   | Anal Date  | Anal Time |
|-----|--------|------------------|----------|---------|---------------|------|------------|-----------|
| 1   | CAL1   | 0                | 1        | 0.146   | 0.13          |      | 05/02/2023 | 16:00     |
| 2   | CAL2   | 1                | 1        | 1.152   | 1.20          | 20.2 | 05/02/2023 | 16:03     |
| 3   | CAL3   | 5                | 1        | 4.384   | 4.65          | -7   | 05/02/2023 | 16:06     |
| 4   | CAL4   | 10               | 1        | 9.480   | 10.09         | 0.9  | 05/02/2023 | 16:10     |
| 5   | CAL5   | 20               | 1        | 18.600  | 19.83         | -0.9 | 05/02/2023 | 16:13     |
| 6   | CAL6   | 40               | 1        | 37.600  | 40.11         | 0.3  | 05/02/2023 | 16:17     |



## Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB125253

| Seq | Lab ID     | True Value<br>(NTU) | Dilution | Reading | Result<br>(NTU) | AnalDate   | Anal Time |
|-----|------------|---------------------|----------|---------|-----------------|------------|-----------|
| 1   | ICV        | 10                  | 1        | 9.380   | 9.984           | 05/02/2023 | 16:20     |
| 2   | ICB        |                     | 1        | 0.142   | 0.124           | 05/02/2023 | 16:23     |
| 3   | CCV1       | 10                  | 1        | 9.330   | 9.931           | 05/02/2023 | 16:27     |
| 4   | CCB1       |                     | 1        | 0.148   | 0.130           | 05/02/2023 | 16:30     |
| 5   | LB125253BL |                     | 1        | 0.152   | 0.134           | 05/02/2023 | 16:33     |
| 6   | O2213-07   |                     | 1        | 0.518   | 0.525           | 05/02/2023 | 16:36     |
| 7   | O2213-08   |                     | 1        | 1.128   | 1.176           | 05/02/2023 | 16:40     |
| 8   | CCV2       | 10                  | 1        | 9.260   | 9.856           | 05/02/2023 | 16:43     |
| 9   | CCB2       |                     | 1        | 0.152   | 0.134           | 05/02/2023 | 16:46     |