



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB125254

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP102259
Turbidity Calibration std, 0NTU	WP102252
Turbidity Calibration std, 20NTU	WP102254
Turbidity Calibration std, 40NTU	WP102253
Turbidity Calibration std, 1NTU	WP102257
Turbidity Calibration std, 5NTU	WP102256
Turbidity LOQ std, 1.0NTU	WP102260
Turbidity Calibration - CCV std, 10 NTU	WP102255
10 NTU Standard 500 ml	W2971

Intercept: 0.026

Slope: 0.9369

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.146	0.13		05/02/2023	16:00
2	CAL2	1	1	1.152	1.20	20.2	05/02/2023	16:03
3	CAL3	5	1	4.384	4.65	-7	05/02/2023	16:06
4	CAL4	10	1	9.480	10.09	0.9	05/02/2023	16:10
5	CAL5	20	1	18.600	19.83	-0.9	05/02/2023	16:13
6	CAL6	40	1	37.600	40.11	0.3	05/02/2023	16:17



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB125254

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.380	9.984	05/02/2023	16:20
2	ICB		1	0.142	0.124	05/02/2023	16:23
3	CCV1	10	1	9.330	9.931	05/02/2023	16:27
4	CCB1		1	0.148	0.130	05/02/2023	16:30
5	LB125254BL		1	0.152	0.134	05/02/2023	16:33
6	O2213-10		1	0.518	0.525	05/02/2023	16:36
7	O2213-11		1	1.128	1.176	05/02/2023	16:40
8	CCV2	10	1	9.260	9.856	05/02/2023	16:43
9	CCB2		1	0.152	0.134	05/02/2023	16:46