


LB125260

SFAM01.1

INSTRUMENT ID: CV



Linear

A= 0.0000e+000

B= 1.8195e-004 slope

C= 3.0059e-002 y-intercept

Rho= 0.9999209

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.055	0.055	135	0.000	135					
0.05	0.050					302	0				
0.20	0.200	0.183	-0.017	842	0.0 %	842					
2.50	2.500	2.506	0.006	13607	0.0 %	13607					
5.00	5.000	4.908	-0.092	26811	0.0 %	26811					
7.50	7.500	7.518	0.018	41154	0.0 %	41154					
10.0	10.000	10.030	0.030	54959	0.0 %	54959					

LB125260

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	135	-		0 SFAM01.1	PPB	5/3/2023 11:46	Std
	0.2 S01	842	-		0.2 SFAM01.1	PPB	5/3/2023 11:48	Std
	2.5 S02	13607	-		2.5 SFAM01.1	PPB	5/3/2023 11:51	Std
	5 S03	26811	-		5 SFAM01.1	PPB	5/3/2023 11:53	Std
	7.5 S04	41154	-		7.5 SFAM01.1	PPB	5/3/2023 11:55	Std
	10 S05	54959	-		10 SFAM01.1	PPB	5/3/2023 11:57	Std
ICV059	ICV059	20782	3.8114 -		SFAM01.1	PPB	5/3/2023 12:03	SMPL
ICB059	ICB059	-43	0.0222 -		SFAM01.1	PPB	5/3/2023 12:05	SMPL
CCV002	CCV002	26821	4.9102 -		SFAM01.1	PPB	5/3/2023 12:07	SMPL
CCB002	CCB002	-64	0.0184 -		SFAM01.1	PPB	5/3/2023 12:10	SMPL
PB152552BL	PBS552	23	0.0342 -		SFAM01.1	PPB	5/3/2023 12:12	SMPL
O2215-03	MDL-SOIL	174	0.0617 -		SFAM01.1	PPB	5/3/2023 12:17	SMPL
PB152553BL	PBS553	104	0.049 -		SFAM01.1	PPB	5/3/2023 12:19	SMPL
O2481-01	MH2869	4345	0.8206 -		SFAM01.1	PPB	5/3/2023 12:21	SMPL
O2481-02	MH2869D	4532	0.8547 -		SFAM01.1	PPB	5/3/2023 12:23	SMPL
O2481-03	MH2869S	15687	2.8843 -		SFAM01.1	PPB	5/3/2023 12:26	SMPL
O2481-04	MH2870	999	0.2118 -		SFAM01.1	PPB	5/3/2023 12:28	SMPL
O2481-05	MH2871	928	0.1989 -		SFAM01.1	PPB	5/3/2023 12:30	SMPL
O2481-06	MH2872	702	0.1578 -		SFAM01.1	PPB	5/3/2023 12:32	SMPL
O2481-07	MH2873	1478	0.299 -		SFAM01.1	PPB	5/3/2023 12:35	SMPL
CCV003	CCV003	26595	4.869 -		SFAM01.1	PPB	5/3/2023 12:37	SMPL
CCB003	CCB003	-83	0.015 -		SFAM01.1	PPB	5/3/2023 12:39	SMPL