

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

5/3/2023 10:36

Reviewed by : 12 Instrument ID : Konelab

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.271	0.0	0.074	
ICB1	0.751	0.0	0.002	
CCV1	245.995	0.0	0.188	
CCB1	0.582	0.0	0.001	
PB152541BL	0.676	0.0	0.001	
O2556-01	0.539	0.0	0.001	
O2556-01DUP	0.458	0.0	0.001	
CCV2	255.216	0.0	0.195	
CCB2	0.410	0.0	0.001	
N	9			
Mean	66.766			
SD	108.8573			
CV%	163.04			

Aquakem v. 7.2AQ1

Results from time period:

Wed May 03 10:06:30 2023

Wed May 03 10:33:17 2023

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.1495	µg/l	5/3/2023 10:06:30	
5.0PPBCN	A	Total CN	P	5.3067	µg/l	5/3/2023 10:06:31	
10PPBCN	A	Total CN	P	10.1826	µg/l	5/3/2023 10:06:32	
50PPBCN	A	Total CN	P	49.395	µg/l	5/3/2023 10:06:33	
100PPBCN	A	Total CN	P	99.4743	µg/l	5/3/2023 10:06:34	
250PPBCN	A	Total CN	P	250.6659	µg/l	5/3/2023 10:06:35	
500PPBCN	A	Total CN	P	499.826	µg/l	5/3/2023 10:06:36	
ICV1	S	Total CN	P	96.2712	µg/l	5/3/2023 10:28:56	
ICB1	S	Total CN	P	0.7508	µg/l	5/3/2023 10:28:57	
CCV1	S	Total CN	P	245.9951	µg/l	5/3/2023 10:29:00	
CCB1	S	Total CN	P	0.5818	µg/l	5/3/2023 10:29:02	
PB152541BL	S	Total CN	P	0.676	µg/l	5/3/2023 10:29:03	
O2556-01	S	Total CN	P	0.5387	µg/l	5/3/2023 10:29:05	
O2556-01DUP	S	Total CN	P	0.4579	µg/l	5/3/2023 10:33:14	
CCV2	S	Total CN	P	255.2163	µg/l	5/3/2023 10:33:16	
CCB2	S	Total CN	P	0.4099	µg/l	5/3/2023 10:33:17	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

5/3/2023 10:07

Reviewed by : 12 Instrument ID : Konelab


Test Total CN

Accepted 5/3/2023 10:07

Factor 1313
Bias 0.001

Coeff. of det. 0.999994

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	0.1495	0.0000	-
2	5.0PPBCN	0.005	5.3067	5.0000	6.1
3	10PPBCN	0.009	10.1826	10.0000	1.8
4	50PPBCN	0.039	49.3950	50.0000	-1.2
5	100PPBCN	0.077	99.4743	100.0000	-0.5
6	250PPBCN	0.192	250.6659	250.0000	0.3
7	500PPBCN	0.382	499.8260	500.0000	2.0

12
5/3/23