



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/4/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 05/03/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 05/04/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %splids-oven

QC:LB125263

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2586-01	50032	15	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O2588-01	ZONE-B-0-3-05012023	1	1.14	8.71	9.85	7.81	76.6	
O2588-02	ZONE-B-3-6-05012023	2	1.16	8.50	9.66	7.49	74.5	
O2588-03	ZONE-B-6-9-05012023	3	1.13	8.59	9.72	7.32	72.1	
O2588-04	ZONE-C-0-3-05012023	4	1.14	8.81	9.95	7.14	68.1	
O2588-05	ZONE-C-3-6-05012023	5	1.14	8.45	9.59	7.02	69.6	
O2588-06	ZONE-C-6-9-05012023	6	1.19	8.67	9.86	7.42	71.9	
O2588-07	ZONE-A-0-3-05022023	7	1.16	8.61	9.77	6.88	66.4	
O2588-08	ZONE-A-3-6-05022023	8	1.19	8.35	9.54	6.56	64.3	
O2588-09	ZONE-A-6-9-05022023	9	1.19	8.75	9.94	7.09	67.4	
O2588-10	ZONE-D-0-3-05022023	10	1.1	8.79	9.89	7.78	76.0	
O2588-11	ZONE-D-3-6-05022023	11	1.19	8.67	9.86	7.37	71.3	
O2588-12	ZONE-D-6-9-05022023	12	1.14	8.78	9.92	7.42	71.5	
O2588-14	PILE-A-UNK-05022023	13	1.15	8.83	9.98	7.34	70.1	
O2588-15	PILE-B-UNK-05022023	14	1.1	8.70	9.8	8.43	84.3	
O2590-01	ETGI-291	16	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O2596-01	XYZ-1502	17	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O2596-02	XYZ-1504	18	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O2596-03	XYZ-1520	19	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O2596-04	XYZ-1526	20	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O2596-05	XYZ-1527	21	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O2596-06	XYZ-1530	22	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O2596-07	XYZ-1531	23	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O2596-08	XYZ-1532	24	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O2596-09	XYZ-1533	25	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O2596-10	XYZ-1536	26	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O2598-01	VNJ-241	27	1.18	8.72	9.9	9.1	90.8	
O2598-03	VNJ-241-EPH-2	28	1.12	8.56	9.68	8.46	85.7	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/4/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 05/03/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 05/04/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %splids-oven

QC:LB125263

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2598-04	ETGI-328	29	1.18	8.38	9.56	8.5	87.4	
O2598-06	ETGI-328-EPH-2	30	1.15	8.80	9.95	9.31	92.7	
O2603-01	BUR-SOILCOMP	31	1.19	8.48	9.67	8.98	91.9	
O2603-02	BUR-SOILCOMP-EPH-2	32	1.13	8.46	9.59	8.86	91.4	
O2604-01	TF-1	33	1.19	8.45	9.64	9.03	92.8	
O2605-01	RC-1	34	1.16	8.62	9.78	8.33	83.2	
O2607-01	HT-1	35	1.11	8.50	9.61	7.75	78.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$