

LB125276


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

46903



A= 0.0000e+000

B= 2.1657e-004

C= 2.4787e-002

Rho= 0.9992922

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/ol)
0.0	0.000	0.061	0.061	169	0.000	169					
0.2	0.200	0.185	-0.015	742	0.0 %	742					
2.5	2.500	2.624	0.124	12000	0.0 %	12000					
5.0	5.000	4.840	-0.160	22233	0.0 %	22233					
7.5	7.500	7.307	-0.193	33624	0.0 %	33624					
10.0	10.000	10.183	0.183	46903	0.0 %	46903					

0/ol)

1-2

1-5

1-3

1-3

2

LB125276 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 May 2023 09:28:35

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	169	-	0.0000	7470A	PPB	04 May 2023 09:55:31	S	Std
0.2 - 1	50.9	742	-	0.2000	7470A	PPB	04 May 2023 09:57:48	S	Std
2.5 - 1	52.5	12000	-	2.5000	7470A	PPB	04 May 2023 10:00:06	S	Std
5.0 - 1	55	22233	-	5.0000	7470A	PPB	04 May 2023 10:02:23	S	Std
7.5 - 1	57.5	33624	-	7.5000	7470A	PPB	04 May 2023 10:04:40	S	Std
10.0 - 1	510	46903	-	10.0000	7470A	PPB	04 May 2023 10:12:06	S	Std
ICV61 - 1	ICV61	17104	3.7291	-	7470A	PPB	04 May 2023 10:15:05	U	SMPL
ICB61 - 1	ICB61	-156	-0.0090	-	7470A	PPB	04 May 2023 10:17:20	U	SMPL
CCV08 - 1	CCV08	22612	4.9219	-	7470A	PPB	04 May 2023 10:19:36	U	SMPL
CCB08 - 1	CCB08	-170	-0.0120	-	7470A	PPB	04 May 2023 10:21:51	U	SMPL
CRA - 1	CRA	814	0.2011	-	7470A	PPB	04 May 2023 10:26:43	U	SMPL
HighStd - 1	HighStd	46227	10.0363	-	7470A	PPB	04 May 2023 10:28:58	U	SMPL
ChkStd - 1	ChkStd	30553	6.6418	-	7470A	PPB	04 May 2023 10:31:14	U	SMPL
PB152550BL - 1	PBW	65	0.0389	-	7470A	PPB	04 May 2023 10:35:34	U	SMPL
PB152550BS - 1	LCSW	16842	3.6723	-	7470A	PPB	04 May 2023 10:37:49	U	SMPL
O2566-01 - 1	AUD-1985	178	0.0633	-	7470A	PPB	04 May 2023 10:40:06	U	SMPL
O2566-02 - 1	AUD-1986	397	0.1108	-	7470A	PPB	04 May 2023 10:42:23	U	SMPL
O2566-03 - 1	AUD-1987	514	0.1361	-	7470A	PPB	04 May 2023 10:44:41	U	SMPL
O2566-04 - 1	AUD-1988	455	0.1233	-	7470A	PPB	04 May 2023 10:46:58	U	SMPL
O2566-05 - 1	AUD-1989	669	0.1697	-	7470A	PPB	04 May 2023 10:49:13	U	SMPL
CCV09 - 1	CCV09	22259	4.8455	-	7470A	PPB	04 May 2023 10:51:29	U	SMPL
CCB09 - 1	CCB09	-146	-0.0068	-	7470A	PPB	04 May 2023 10:53:45	U	SMPL
O2566-06 - 1	AUD-1990	583	0.1510	-	7470A	PPB	04 May 2023 10:56:02	U	SMPL
O2566-07 - 1	AUD-1991	539	0.1415	-	7470A	PPB	04 May 2023 10:58:18	U	SMPL
O2566-08 - 1	AUD-1992	361	0.1030	-	7470A	PPB	04 May 2023 11:00:34	U	SMPL
O2566-09 - 1	AUD-1993	546	0.1430	-	7470A	PPB	04 May 2023 11:02:50	U	SMPL
O2566-09DUP - 1	AUD-1993DUP	536	0.1409	-	7470A	PPB	04 May 2023 11:05:06	U	SMPL
O2566-09MS - 1	AUD-1993MS	16331	3.5616	-	7470A	PPB	04 May 2023 11:07:23	U	SMPL
O2566-09MSD - 1	AUD-1993MSD	16105	3.5127	-	7470A	PPB	04 May 2023 11:09:39	U	SMPL
PB152551BL - 1	PBW	-170	-0.0120	-	7470A	PPB	04 May 2023 11:11:56	U	SMPL
PB152551BS - 1	LCSW	17227	3.7557	-	7470A	PPB	04 May 2023 11:16:50	U	SMPL
O2213-07 - 1	LOD-MDL-WATER-01-QT2-2023	185	0.0649	-	7470A	PPB	04 May 2023 11:23:36	U	SMPL
CCV10 - 1	CCV10	22283	4.8507	-	7470A	PPB	04 May 2023 11:25:51	U	SMPL
CCB10 - 1	CCB10	-145	-0.0066	-	7470A	PPB	04 May 2023 11:28:07	U	SMPL
O2213-08 - 1	LOQ-WATER-02-QT2-2023	812	0.2006	-	7470A	PPB	04 May 2023 11:30:24	U	SMPL
O2552-01 - 1	SAYR-1120	3876	0.8642	-	7470A	PPB	04 May 2023 11:32:40	U	SMPL
O2552-03 - 1	SAYR-1121	44035	9.5616	-	7470A	PPB	04 May 2023 11:34:56	U	SMPL
O2556-03 - 1	MW-1-4-WATER	170	0.0616	-	7470A	PPB	04 May 2023 11:37:11	U	SMPL
O2567-03 - 1	EB-2023-29-06	431	0.1181	-	7470A	PPB	04 May 2023 11:39:29	U	SMPL
O2581-01 - 1	10868522	877	0.2147	-	7470A	PPB	04 May 2023 11:41:46	U	SMPL
O2581-01DUP - 1	10868522DUP	784	0.1946	-	7470A	PPB	04 May 2023 11:44:02	U	SMPL
O2581-01MS - 1	10868522MS	16609	3.6218	-	7470A	PPB	04 May 2023 11:46:19	U	SMPL
O2581-01MSD - 1	10868522MSD	17583	3.8328	-	7470A	PPB	04 May 2023 11:51:38	U	SMPL
CCV11 - 1	CCV11	22083	4.8074	-	7470A	PPB	04 May 2023 11:53:53	U	SMPL
CCB11 - 1	CCB11	-59	0.0120	-	7470A	PPB	04 May 2023 11:56:10	U	SMPL
PB152532TB - 1	PB152532TB	61	0.0380	-	7470A	PPB	04 May 2023 12:01:37	U	SMPL
O2566-09LX5 - 1		182	0.0642	-	7470A	PPB	04 May 2023 12:03:52	U	SMPL
O2566-09A - 1		16889	3.6825	-	7470A	PPB	04 May 2023 12:06:08	U	SMPL
O2581-01LX5 - 1		70	0.0396	-	7470A	PPB	04 May 2023 12:08:24	U	SMPL
O2581-01A - 1		18105	3.9459	-	7470A	PPB	04 May 2023 12:10:40	U	SMPL
CCV12 - 1	CCV12	22446	4.8860	-	7470A	PPB	04 May 2023 12:17:55	U	SMPL
CCB12 - 1	CCB12	-67	0.0103	-	7470A	PPB	04 May 2023 12:20:12	U	SMPL