

LB125285


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

60923



A= 0.0000e+000

B= 1.6606e-004

C= -1.9470e-002

Rho= 0.9996153

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.090	0.090	657	0.000	657					
0.2	0.200	0.199	-0.001	1314	0.0 %	1314					1
2.5	2.500	2.534	0.034	15375	0.0 %	15375					1
5.0	5.000	4.790	-0.210	28962	0.0 %	28962					4
7.5	7.500	7.491	-0.009	45228	0.0 %	45228					0
10.0	10.000	10.097	0.097	60923	0.0 %	60923					1

LB125285 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 May 2023 15:00:01

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	657	-	0.0007471B	PPB		04 May 2023 15:04:55	S	Std
0.2 - 1	50.2	1314	-	0.20007471B	PPB		04 May 2023 15:07:13	S	Std
2.5 - 1	52.1	15375	-	2.50007471B	PPB		04 May 2023 15:09:30	S	Std
5.0 - 1	55	28962	-	5.00007471B	PPB		04 May 2023 15:11:48	S	Std
7.5 - 1	57.1	45228	-	7.50007471B	PPB		04 May 2023 15:16:41	S	Std
10.0 - 1	516	60923	-	10.00007471B	PPB		04 May 2023 15:22:13	S	Std
ICV62 - 1	ICV62	22684	3.7474	-7471B	PPB		04 May 2023 15:26:02	U	SMPL
ICB62 - 1	ICB62	435	0.0528	-7471B	PPB		04 May 2023 15:28:17	U	SMPL
CCV13 - 1	CCV13	29627	4.9003	-7471B	PPB		04 May 2023 15:30:35	U	SMPL
CCB13 - 1	CCB13	391	0.0455	-7471B	PPB		04 May 2023 15:32:50	U	SMPL
CRA - 1	CRA	1275	0.1923	-7471B	PPB		04 May 2023 15:35:08	U	SMPL
HighStd - 1	HighStd	61268	10.1545	-7471B	PPB		04 May 2023 15:37:24	U	SMPL
ChkStd - 1	ChkStd	43746	7.2448	-7471B	PPB		04 May 2023 15:49:10	U	SMPL
PB152562BL - 1	PBS	352	0.0390	-7471B	PPB		04 May 2023 15:51:25	U	SMPL
PB152562BS - 1	LCSS	23663	3.9099	-7471B	PPB		04 May 2023 15:56:16	U	SMPL
O2590-01 - 1	ETGI-291	18539	3.0591	-7471B	PPB		04 May 2023 16:00:51	U	SMPL
O2598-01 - 1	VNI-241	6491	1.0584	-7471B	PPB		04 May 2023 16:03:08	U	SMPL
O2598-04 - 1	ETGI-328	9802	1.6082	-7471B	PPB		04 May 2023 16:05:27	U	SMPL
O2603-01 - 1	BUR-SOILCOMP	10081	1.6545	-7471B	PPB		04 May 2023 16:07:43	U	SMPL
CCV14 - 1	CCV14	29242	4.8364	-7471B	PPB		04 May 2023 16:09:58	U	SMPL
CCB14 - 1	CCB14	472	0.0588	-7471B	PPB		04 May 2023 16:12:15	U	SMPL
O2604-01 - 1	TF-1	2131	0.3344	-7471B	PPB		04 May 2023 16:14:33	U	SMPL
O2604-01DUP - 1	TF-1DUP	2095	0.3284	-7471B	PPB		04 May 2023 16:16:54	U	SMPL
O2604-01MS - 1	TF-1MS	26625	4.4018	-7471B	PPB		04 May 2023 16:19:10	U	SMPL
O2604-01MSD - 1	TF-1MSD	25902	4.2817	-7471B	PPB		04 May 2023 16:21:26	U	SMPL
O2605-01 - 1	RC-1	6133	0.9990	-7471B	PPB		04 May 2023 16:23:44	U	SMPL
O2607-01 - 1	HT-1	3413	0.5473	-7471B	PPB		04 May 2023 16:26:02	U	SMPL
PB152563BL - 1	PBS	627	0.0846	-7471B	PPB		04 May 2023 16:28:19	U	SMPL
PB152563BS - 1	LCSS	24042	3.9729	-7471B	PPB		04 May 2023 16:33:22	U	SMPL
O2588-01 - 1	ZONE-B-0-3-05012023	20062	3.3120	-7471B	PPB		04 May 2023 16:40:41	U	SMPL
O2588-01DUP - 1	ZONE-B-0-3-05012023DUP	20128	3.3229	-7471B	PPB		04 May 2023 16:42:57	U	SMPL
CCV15 - 1	CCV15	29589	4.8940	-7471B	PPB		04 May 2023 16:45:14	U	SMPL
CCB15 - 1	CCB15	469	0.0584	-7471B	PPB		04 May 2023 16:47:30	U	SMPL
O2588-01MS - 1	ZONE-B-0-3-05012023MS	42233	6.9936	-7471B	PPB		04 May 2023 17:01:33	U	SMPL
O2588-01MSD - 1	ZONE-B-0-3-05012023MSD	42632	7.0599	-7471B	PPB		04 May 2023 17:06:30	U	SMPL
O2588-02 - 1	ZONE-B-3-6-05012023	8556	1.4013	-7471B	PPB		04 May 2023 17:08:47	U	SMPL
O2588-03 - 1	ZONE-B-6-9-05012023	12670	2.0845	-7471B	PPB		04 May 2023 17:11:05	U	SMPL
O2588-04 - 1	ZONE-C-0-3-05012023	9659	1.5845	-7471B	PPB		04 May 2023 17:13:22	U	SMPL
O2588-05 - 1	ZONE-C-3-6-05012023	11748	1.9314	-7471B	PPB		04 May 2023 17:15:40	U	SMPL
O2588-06 - 1	ZONE-C-6-9-05012023	17141	2.8269	-7471B	PPB		04 May 2023 17:17:58	U	SMPL
O2588-07 - 1	ZONE-A-0-3-05022023	10396	1.7069	-7471B	PPB		04 May 2023 17:20:15	U	SMPL
O2588-08 - 1	ZONE-A-3-6-05022023	30419	5.0318	-7471B	PPB		04 May 2023 17:22:34	U	SMPL
O2588-09 - 1	ZONE-A-6-9-05022023	10384	1.7049	-7471B	PPB		04 May 2023 17:24:50	U	SMPL
CCV16 - 1	CCV16	29636	4.9018	-7471B	PPB		04 May 2023 17:27:08	U	SMPL
CCB16 - 1	CCB16	513	0.0657	-7471B	PPB		04 May 2023 17:29:24	U	SMPL
O2588-10 - 1	ZONE-D-0-3-05022023	12698	2.0891	-7471B	PPB		04 May 2023 17:31:41	U	SMPL
O2588-11 - 1	ZONE-D-3-6-05022023	12650	2.0811	-7471B	PPB		04 May 2023 17:33:57	U	SMPL
O2588-12 - 1	ZONE-D-6-9-05022023	14360	2.3651	-7471B	PPB		04 May 2023 17:36:15	U	SMPL
O2588-14 - 1	PILE-A-UNK-05022023	10886	1.7882	-7471B	PPB		04 May 2023 17:38:31	U	SMPL
O2588-15 - 1	PILE-B-UNK-05022023	6144	1.0008	-7471B	PPB		04 May 2023 17:40:49	U	SMPL
O2213-01 - 1	LOD-MDL-SOIL-01-QT2-2023	408	0.0483	-7471B	PPB		04 May 2023 17:48:08	U	SMPL
O2213-01RE - 1	LOD-MDL-SOIL-01-QT2-2023RE	1024	0.1506	-7471B	PPB		04 May 2023 17:50:25	U	SMPL
O2213-03 - 1	MDL-SOIL-03-QT2-2023	648	0.0881	-7471B	PPB		04 May 2023 17:52:42	U	SMPL
CCV17 - 1	CCV17	29306	4.8470	-7471B	PPB		04 May 2023 17:54:59	U	SMPL
CCB17 - 1	CCB17	111	-0.0010	-7471B	PPB		04 May 2023 17:57:15	U	SMPL
O2604-01LX5 - 1		367	0.0415	-7471B	PPB		04 May 2023 17:59:33	U	SMPL
O2604-01A - 1		23511	3.8847	-7471B	PPB		04 May 2023 18:01:48	U	SMPL
O2588-01LX5 - 1		3872	0.6235	-7471B	PPB		04 May 2023 18:04:04	U	SMPL
O2588-01A - 1		48912	8.1027	-7471B	PPB		04 May 2023 18:06:22	U	SMPL
CCV18 - 1	CCV18	29021	4.7997	-7471B	PPB		04 May 2023 18:08:38	U	SMPL
CCB18 - 1	CCB18	20	-0.0161	-7471B	PPB		04 May 2023 18:10:56	U	SMPL