



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB125295

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	05/05/2023	09:10
2	CAL2	1	Water	NA	NA	20.6	7.00	05/05/2023	09:11
3	CAL3	1	Water	NA	NA	20.6	10.02	05/05/2023	09:12
4	ICV	1	Water	NA	NA	20.7	6.98	05/05/2023	09:15
5	CCV1	1	Water	NA	NA	20.6	2.01	05/05/2023	09:20
6	O2581-01	1	Water	NA	NA	20.6	6.63	05/05/2023	08:22
7	O2581-01DUP	1	Water	NA	NA	20.7	6.64	05/05/2023	09:23
8	CCV2	1	Water	NA	NA	20.6	12.02	05/05/2023	09:25

WORKLIST(Hardcopy Internal Chain)

03 125295

WorkList Name : ph 02632		WorkList ID : 169688		Department : Wet-Chemistry		Date : 05-05-2023 08:51:32	
Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample
05/02/2023	Water	Q2581-01	pH				
				Cool 4 deg C	PSEG03	N11	10868522
						05/02/2023	9040C

Date/Time 05/05/23 09:06
 Raw Sample Received by: JDCSM
 Raw Sample Relinquished by: JDCSM

Date/Time 05/05/23 10:00
 Raw Sample Received by: JDCSM
 Raw Sample Relinquished by: JDCSM