



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB125298

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	05/05/2023	10:20
2	CAL2	1	Water	NA	NA	20.7	7.00	05/05/2023	10:21
3	CAL3	1	Water	NA	NA	20.7	10.02	05/05/2023	10:22
4	ICV	1	Water	NA	NA	20.6	7.00	05/05/2023	10:25
5	CCV1	1	Water	NA	NA	20.8	2.01	05/05/2023	10:30
6	O2598-02	1	Solid	20.02	20	21.5	9.43	05/05/2023	10:32
7	O2598-02DUP	1	Solid	20.03	20	21.6	9.44	05/05/2023	10:33
8	O2598-05	1	Solid	20.05	20	21.1	10.38	05/05/2023	10:40
9	O2611-01	1	Solid	20.02	20	21.5	6.83	05/05/2023	10:42
10	O2624-06	1	Solid	20.02	20	21.9	6.50	05/05/2023	10:47
11	O2624-07	1	Solid	20.03	20	21.7	7.10	05/05/2023	10:50
12	O2624-08	1	Solid	20.01	20	21.8	8.91	05/05/2023	11:00
13	O2626-01	1	Solid	20.02	20	21.5	6.84	05/05/2023	11:10
14	CCV2	1	Water	NA	NA	20.6	12.02	05/05/2023	11:15

UB 125298

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 02598

WorkList ID : 169704

Department : Wet-Chemistry

Date : 05-05-2023 10:11:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/03/2023	Solid	O2598-02	Corrosivity	Cool 4 deg C	PSEG03	O11	VNJ-241	05/03/2023	9045D
05/03/2023	Solid	O2598-05	Corrosivity	Cool 4 deg C	PSEG03	O11	ETGI-328	05/03/2023	9045D
05/04/2023	Solid	O2611-01	Corrosivity	Cool 4 deg C	PSEG03	O11	357	05/04/2023	9045D
05/04/2023	Solid	O2624-06	Corrosivity	Cool 4 deg C	EARTH03	O11	WZ-3	05/04/2023	9045D
05/04/2023	Solid	O2624-07	Corrosivity	Cool 4 deg C	EARTH03	O11	WZ-2A	05/04/2023	9045D
05/04/2023	Solid	O2624-08	Corrosivity	Cool 4 deg C	EARTH03	O11	WZ-2B	05/04/2023	9045D
05/04/2023	Solid	O2626-01	Corrosivity	Cool 4 deg C	PSEG03	O11	D2978	05/04/2023	9045D

Date/Time 05/05/23 10:16

Raw Sample Received by: JP (WCS)

Raw Sample Relinquished by: JP (WCS)

Date/Time 05/05/23 13:00

Raw Sample Received by: JP (WCS)

Raw Sample Relinquished by: JP (WCS)