


7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.1109e-004 slope

C = -5.9868×10^{-3} $\gamma = 1.5 \times 10^{-4}$

Rho= 0.9990138

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	0.011	0.011	81	0.000	81	0				17
0.2	0.200	0.233	0.033	1132	0.0 %	1132					17
2.5	2.500	2.263	-0.237	10748	0.0 %	10748					9
5.0	5.000	5.304	0.304	25156	0.0 %	25156					6
7.5	7.500	7.494	-0.006	35530	0.0 %	35530					6
10.0	10.000	9.895	-0.105	46904	0.0 %	46904					0

LB125302 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 May 2023 11:19:47

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	81	-	0.0000	7470A	PPB	05 May 2023 11:22:43	S	Std
0.2 - 1	50.2	1132	-	0.2000	7470A	PPB	05 May 2023 11:32:53	S	Std
2.5 - 1	52.5	10748	-	2.5000	7470A	PPB	05 May 2023 11:35:11	S	Std
5.0 - 1	55	25156	-	5.0000	7470A	PPB	05 May 2023 11:37:29	S	Std
7.5 - 1	57.5	35530	-	7.5000	7470A	PPB	05 May 2023 11:39:45	S	Std
10.0 - 1	57.0	46904	-	10.0000	7470A	PPB	05 May 2023 11:42:03	S	Std
ICV63 - 1	ICV63	18610	3.9224	-7470A	PPB	PPB	05 May 2023 11:44:49	U	SMPL
ICB63 - 1	ICB63	-134	-0.0343	-7470A	PPB	PPB	05 May 2023 11:47:04	U	SMPL
CCV19 - 1	CCV19	25160	5.3050	-7470A	PPB	PPB	05 May 2023 11:49:21	U	SMPL
CCB19 - 1	CCB19	-149	-0.0374	-7470A	PPB	PPB	05 May 2023 11:51:36	U	SMPL
CRA - 1	CRA	1091	0.2243	-7470A	PPB	PPB	05 May 2023 11:53:53	U	SMPL
HighStd - 1	HighStd	46721	9.8563	-7470A	PPB	PPB	05 May 2023 11:56:09	U	SMPL
ChkStd - 1	ChkStd	34106	7.1934	-7470A	PPB	PPB	05 May 2023 11:58:24	U	SMPL
PB152586BL - 1	PBW	-254	-0.0596	-7470A	PPB	PPB	05 May 2023 12:00:42	U	SMPL
PB152586BS - 1	LCSW	18033	3.8006	-7470A	PPB	PPB	05 May 2023 12:05:34	U	SMPL
O2590-02 - 1	ETGI-291	-122	-0.0317	-7470A	PPB	PPB	05 May 2023 12:07:50	U	SMPL
O2598-02 - 1	VNJ-241	18	-0.0022	-7470A	PPB	PPB	05 May 2023 12:10:08	U	SMPL
O2598-05 - 1	ETGI-328	67	0.0082	-7470A	PPB	PPB	05 May 2023 12:12:25	U	SMPL
O2598-05DUP - 1	ETGI-328DUP	111	0.0174	-7470A	PPB	PPB	05 May 2023 12:14:43	U	SMPL
O2598-05MS - 1	ETGI-328MS	20204	4.2589	-7470A	PPB	PPB	05 May 2023 12:16:58	U	SMPL
CCV20 - 1	CCV20	24981	5.2672	-7470A	PPB	PPB	05 May 2023 12:19:14	U	SMPL
CCB20 - 1	CCB20	-148	-0.0372	-7470A	PPB	PPB	05 May 2023 12:21:31	U	SMPL
O2598-05MSD - 1	ETGI-328MSD	18626	3.9258	-7470A	PPB	PPB	05 May 2023 12:23:49	U	SMPL
PB152614BL - 1	PBW	-81	-0.0231	-7470A	PPB	PPB	05 May 2023 12:26:04	U	SMPL
PB152614BS - 1	LCSW	20182	4.2542	-7470A	PPB	PPB	05 May 2023 12:28:22	U	SMPL
O2213-07 - 1	LOD-MDI-WATER-01QT2-2023	248	0.0464	-7470A	PPB	PPB	05 May 2023 12:33:12	U	SMPL
O2213-07RE - 1	LOD-MDI-WATER-01QT2-2023RE	802	0.1633	-7470A	PPB	PPB	05 May 2023 12:40:51	U	SMPL
O2213-09 - 1	MDL-WATER-03-QT2-2023	280	0.0531	-7470A	PPB	PPB	05 May 2023 12:43:07	U	SMPL
O2587-02 - 1	COMP	352	0.0683	-7470A	PPB	PPB	05 May 2023 12:45:24	U	SMPL
O2587-02DUP - 1	COMPDUP	381	0.0744	-7470A	PPB	PPB	05 May 2023 12:50:19	U	SMPL
O2587-02MS - 1	COMPMS	16695	3.5181	-7470A	PPB	PPB	05 May 2023 12:52:36	U	SMPL
O2587-02MSD - 1	COMPMSD	17819	3.7554	-7470A	PPB	PPB	05 May 2023 12:54:52	U	SMPL
CCV21 - 1	CCV21	25167	5.3065	-7470A	PPB	PPB	05 May 2023 12:57:10	U	SMPL
CCB21 - 1	CCB21	-149	-0.0374	-7470A	PPB	PPB	05 May 2023 12:59:27	U	SMPL
PB152561TB - 1	PB152561TB	11	-0.0037	-7470A	PPB	PPB	05 May 2023 13:01:46	U	SMPL
O2598-05LX5 - 1		55	0.0056	-7470A	PPB	PPB	05 May 2023 13:04:02	U	SMPL
O2598-05A - 1		19469	4.1037	-7470A	PPB	PPB	05 May 2023 13:06:18	U	SMPL
O2587-02LX5 - 1		-42	-0.0149	-7470A	PPB	PPB	05 May 2023 13:08:34	U	SMPL
O2587-02A - 1		19275	4.0628	-7470A	PPB	PPB	05 May 2023 13:10:51	U	SMPL
CCV22 - 1	CCV22	25254	5.3249	-7470A	PPB	PPB	05 May 2023 13:13:08	U	SMPL
CCB22 - 1	CCB22	-432	-0.0972	-7470A	PPB	PPB	05 May 2023 13:15:26	U	SMPL