

LB125307


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49134



A= 0.0000e+000

B= 2.0350e-004 slope

C= -1.2096e-002 intercept

Rho= 0.9999792

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.006	0.006	87	0.000	87					11
0.05	0.050					302	0				11
0.20	0.200	0.214	0.014	1109	0.0 %	1109					11
2.50	2.500	2.480	-0.020	12245	0.0 %	12245					11
5.00	5.000	4.972	-0.028	24492	0.0 %	24492					11
7.50	7.500	7.543	0.043	37124	0.0 %	37124					11
10.0	10.000	9.987	-0.013	49134	0.0 %	49134					0

LB125307

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	87	-		0 SFAM01.1	PPB	5/5/2023 14:04	Std
	0.2 S01	1109	-		0.2 SFAM01.1	PPB	5/5/2023 14:06	Std
	2.5 S02	12245	-		2.5 SFAM01.1	PPB	5/5/2023 14:08	Std
	5 S03	24492	-		5 SFAM01.1	PPB	5/5/2023 14:10	Std
	7.5 S04	37124	-		7.5 SFAM01.1	PPB	5/5/2023 14:13	Std
	10 S05	49134	-		10 SFAM01.1	PPB	5/5/2023 14:18	Std
ICV064	ICV064	20360	4.1311 -		SFAM01.1	PPB	5/5/2023 14:20	SMPL
ICB064	ICB064	-286	-0.0703 -		SFAM01.1	PPB	5/5/2023 14:23	SMPL
CCV023	CCV023	24709	5.0161 -		SFAM01.1	PPB	5/5/2023 14:25	SMPL
CCB023	CCB023	-225	-0.0579 -		SFAM01.1	PPB	5/5/2023 14:27	SMPL
PB152617BL	PBW617	0	-0.0121 -		SFAM01.1	PPB	5/5/2023 14:30	SMPL
O2215-01	MDL-WATER	275	0.0439 -		SFAM01.1	PPB	5/5/2023 14:32	SMPL
CCV024	CCV024	24573	4.9885 -		SFAM01.1	PPB	5/5/2023 14:34	SMPL
CCB024	CCB024	-151	-0.0428 -		SFAM01.1	PPB	5/5/2023 14:36	SMPL