

LB125312


7471B

INSTRUMENT ID: CV2

Linear

μ Abs.:

6346E



A= 0.0000e+000

B= 1.5671e-004

C= 5.5057e-002

Rho= 0.9998704

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b
0.0	0.000	0.077	0.077	139	0.000	139					
0.2	0.200	0.194	-0.006	886	0.0 %	886					
2.5	2.500	2.382	-0.118	14848	0.0 %	14848					
5.0	5.000	5.033	0.033	31764	0.0 %	31764					
7.5	7.500	7.514	0.014	47598	0.0 %	47598					
10.0	10.000	10.001	0.001	63466	0.0 %	63466					

0.0
0.2
2.5
5.0
7.5
10.0

LB125312 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 May 2023 15:56:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	139	-	0.0000	7471B	PPB	05 May 2023 16:08:43	S	Std
0.2 - 1	50.2	886	-	0.2000	7471B	PPB	05 May 2023 16:13:32	S	Std
2.5 - 1	52.5	14848	-	2.5000	7471B	PPB	05 May 2023 16:15:49	S	Std
5.0 - 1	55	31764	-	5.0000	7471B	PPB	05 May 2023 16:18:06	S	Std
7.5 - 1	57.5	47598	-	7.5000	7471B	PPB	05 May 2023 16:20:25	S	Std
10.0 - 1	510	63466	-	10.0000	7471B	PPB	05 May 2023 16:22:43	S	Std
ICV65 - 1	ICV65	25133	3.9936	-	7471B	PPB	05 May 2023 16:25:29	U	SMPL
ICB65 - 1	ICB65	-87	0.0414	-	7471B	PPB	05 May 2023 16:27:44	U	SMPL
CCV25 - 1	CCV25	31613	5.0091	-	7471B	PPB	05 May 2023 16:30:02	U	SMPL
CCB25 - 1	CCB25	-82	0.0422	-	7471B	PPB	05 May 2023 16:32:17	U	SMPL
CRA - 1	CRA	902	0.1964	-	7471B	PPB	05 May 2023 16:37:25	U	SMPL
HighStd - 1	HighStd	63536	10.0116	-	7471B	PPB	05 May 2023 16:39:41	U	SMPL
ChkStd - 1	ChkStd	45488	7.1834	-	7471B	PPB	05 May 2023 16:41:56	U	SMPL
PB152604BL - 1	PBS	-196	0.0243	-	7471B	PPB	05 May 2023 16:44:15	U	SMPL
PB152604BS - 1	LCSS	22654	3.6051	-	7471B	PPB	05 May 2023 16:46:33	U	SMPL
O2608-01 - 1	VNJ-227	2358	0.4246	-	7471B	PPB	05 May 2023 16:48:50	U	SMPL
O2616-01 - 1	SP-1	9104	1.4817	-	7471B	PPB	05 May 2023 16:51:08	U	SMPL
O2622-01 - 1	2264	3910	0.6678	-	7471B	PPB	05 May 2023 16:53:25	U	SMPL
O2622-03 - 1	RBR-251675	60266	9.4992	-	7471B	PPB	05 May 2023 16:55:43	U	SMPL
O2622-03DUP - 1	RBR-251675DUP	57816	9.1153	-	7471B	PPB	05 May 2023 16:57:58	U	SMPL
CCV26 - 1	CCV26	31337	4.9658	-	7471B	PPB	05 May 2023 17:00:16	U	SMPL
CCB26 - 1	CCB26	-74	0.0435	-	7471B	PPB	05 May 2023 17:02:34	U	SMPL
O2622-03MSD - 1	RBR-251675MSD	86465	13.6048	-	7471B	PPB	05 May 2023 17:07:07	U	SMPL
O2622-03MS - 1	RBR-251675MS	87121	13.7076	-	7471B	PPB	05 May 2023 17:09:51	U	SMPL
O2624-01 - 1	WZ-3	7721	1.2650	-	7471B	PPB	05 May 2023 17:12:56	U	SMPL
O2624-02 - 1	WZ-2A	24707	3.9268	-	7471B	PPB	05 May 2023 17:15:11	U	SMPL
O2624-03 - 1	WZ-2B	1824	0.3409	-	7471B	PPB	05 May 2023 17:17:27	U	SMPL
O2622-03LX5 - 1		11476	1.8534	-	7471B	PPB	05 May 2023 17:19:46	U	SMPL
O2622-03A - 1		75778	11.9300	-	7471B	PPB	05 May 2023 17:22:03	U	SMPL
CCV27 - 1	CCV27	31497	4.9909	-	7471B	PPB	05 May 2023 17:24:21	U	SMPL
CCB27 - 1	CCB27	-77	0.0430	-	7471B	PPB	05 May 2023 17:26:41	U	SMPL
O2622-03DLX2 - 1		31164	4.9387	-	7471B	PPB	05 May 2023 17:28:59	U	SMPL
O2622-03DUPDLX2 - 1		30100	4.7720	-	7471B	PPB	05 May 2023 17:31:14	U	SMPL
O2622-03MSDLX2 - 1		43302	6.8408	-	7471B	PPB	05 May 2023 17:33:32	U	SMPL
O2622-03MSDDLX2 - 1		43330	6.8452	-	7471B	PPB	05 May 2023 17:35:50	U	SMPL
O2622-03LDLX10 - 1		4608	0.7772	-	7471B	PPB	05 May 2023 17:38:08	U	SMPL
O2622-03ADLX2 - 1		44283	6.9945	-	7471B	PPB	05 May 2023 17:40:26	U	SMPL
CCV28 - 1	CCV28	32072	5.0810	-	7471B	PPB	05 May 2023 17:42:42	U	SMPL
CCB28 - 1	CCB28	-113	0.0373	-	7471B	PPB	05 May 2023 17:45:01	U	SMPL