



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB125319

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	05/05/2023	16:15
2	CAL2	1	Water	NA	NA	20.6	7.00	05/05/2023	16:16
3	CAL3	1	Water	NA	NA	20.7	10.02	05/05/2023	16:19
4	ICV	1	Water	NA	NA	20.7	6.98	05/05/2023	16:20
5	CCV1	1	Water	NA	NA	20.6	2.01	05/05/2023	16:22
6	O2632-03	1	Water	NA	NA	20.7	5.20	05/05/2023	16:25
7	O2642-02	1	Water	NA	NA	20.1	6.98	05/05/2023	16:29
8	O2642-03	1	Water	NA	NA	20.7	8.02	05/05/2023	16:35
9	O2645-01	1	Water	NA	NA	20.7	6.96	05/05/2023	16:40
10	O2648-02	1	Water	NA	NA	20.2	7.36	05/05/2023	16:42
11	O2648-03	1	Water	NA	NA	20.7	6.44	05/05/2023	16:45
12	O2648-04	1	Water	NA	NA	20.3	5.80	05/05/2023	16:47
13	O2648-05	1	Water	NA	NA	20.2	9.29	05/05/2023	16:50
14	O2658-03	1	Water	NA	NA	20.5	4.77	05/05/2023	17:00
15	O2658-04	1	Water	NA	NA	20.5	5.31	05/05/2023	17:02
16	CCV2	1	Water	NA	NA	20.8	12.02	05/05/2023	17:05
17	O2658-05	1	Water	NA	NA	20.1	6.96	05/05/2023	17:10
18	O2658-06	1	Water	NA	NA	20.5	9.01	05/05/2023	17:15
19	O2658-07	1	Water	NA	NA	20.7	4.72	05/05/2023	17:20
20	O2660-01	1	Water	NA	NA	20.1	7.22	05/05/2023	17:25

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	O2660-02	1	Water	NA	NA	20.1	7.47	05/05/2023	17:29
22	O2661-01	1	Water	NA	NA	20.7	3.79	05/05/2023	17:30
23	O2661-02	1	Water	NA	NA	20.6	3.83	05/05/2023	17:33
24	O2661-02DUP	1	Water	NA	NA	20.7	3.84	05/05/2023	17:34
25	CCV3	1	Water	NA	NA	20.7	2.02	05/05/2023	17:36

125319

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O2648

WorkList ID : 169709

Department : Wet-Chemistry

Date : 05-05-2023 14:38:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/05/2023	Water	O2632-03	pH	A	Cool 4 deg C	PSEG03	O11	34336	05/05/2023 9040C
05/04/2023	Water	O2642-02	pH	C	Cool 4 deg C	TETR06	N11	TT-052-IDWGW-20230504	05/04/2023 9040C
05/04/2023	Water	O2642-03	pH	C	Cool 4 deg C	TETR06	N11	TT-053-IDWGW-20230504	05/04/2023 9040C
05/04/2023	Water	O2645-01	pH	A	Cool 4 deg C	STER05	O11	POOL-WATER	05/04/2023 9040C
05/05/2023	Water	O2648-02	pH	C	Cool 4 deg C	PSEG03	O11	ORA-1635	05/05/2023 9040C
05/05/2023	Water	O2648-03	pH	C	Cool 4 deg C	PSEG03	O11	ORA-1636	05/05/2023 9040C
05/05/2023	Water	O2648-04	pH	C	Cool 4 deg C	PSEG03	O11	ORA-1637	05/05/2023 9040C
05/05/2023	Water	O2648-05	pH	A	Cool 4 deg C	PSEG03	O11	ORA-1638	05/05/2023 9040C
05/05/2023	Water	O2658-03	pH	A	Cool 4 deg C	PSEG03	O11	327	05/05/2023 9040C
05/05/2023	Water	O2658-04	pH	A	Cool 4 deg C	PSEG03	O11	328	05/05/2023 9040C
05/05/2023	Water	O2658-05	pH	C	Cool 4 deg C	PSEG03	O11	507	05/05/2023 9040C
05/05/2023	Water	O2658-06	pH	A	Cool 4 deg C	PSEG03	O11	508	05/05/2023 9040C
05/05/2023	Water	O2658-07	pH	C	Cool 4 deg C	PSEG03	O11	509	05/05/2023 9040C
05/05/2023	Water	O2660-01	pH	F	Cool 4 deg C	ENV170	O11	WATER-SAMPLE-API-3-5	05/05/2023 9040C
05/05/2023	Water	O2660-02	pH	F	Cool 4 deg C	ENV170	O11	WATER-SAMPLE-API-3-6	05/05/2023 9040C
04/17/2023	Water	O2661-01	pH	A	Cool 4 deg C	STER14	O11	MINI-CELL	04/17/2023 9040C
04/17/2023	Water	O2661-02	pH	A	Cool 4 deg C	STER14	O11	JS10000	04/17/2023 9040C

Date/Time 05/05/23 16:10

Raw Sample Received by: 70(WC)

Raw Sample Relinquished by: LC (3m)

Date/Time 05/05/23

Raw Sample Received by: 18:00

Raw Sample Relinquished by: 70(WC)