



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125325

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	05/08/2023	09:00
2	O2500-06		1	11	64.00	147.2	147.2	05/08/2023	09:30
3	O2630-01		1	12	100.00	>212.0	>212.0	05/08/2023	10:00
4	O2648-01		1	8	18.00	64.4	64.4	05/08/2023	10:30
5	O2648-01DUP		1	8	18.00	64.4	64.4	05/08/2023	11:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16125325

WorkList Name : f p s 5-8

WorkList ID : 169731

Department : Wet-Chemistry

Date : 05-08-2023 08:24:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/08/2023	Solid	O2500-06	Flash Point	Cool 4 deg C	CHEM02	QA Of	PT-FP-SOIL		
05/11/2023	Solid	O2630-01	Flash Point	Cool 4 deg C	PSEG03	O11	2821	04/24/2023	1010B
05/12/2023	Solid	O2648-01	Flash Point	Cool 4 deg C	PSEG03	O11	912	05/04/2023	1010B
								05/05/2023	1010B

Date/Time 05/08/2023 08:50
 Raw Sample Received by: RMA CWA
 Raw Sample Relinquished by: SP(WC)

Date/Time 05/08/2023 11:40
 Raw Sample Received by: SP(WC)
 Raw Sample Relinquished by: RMA CWA