


LB125327

SFAM01.1

INSTRUMENT ID: CY1

Linear



A= 0.0000e+000
 B= 2.0116e-004 slope
 C= -1.1608e-002 y-intercept
 Rho= 0.9999562
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.031	0.031	211	0.000	211					
0.05	0.050					302	0				
0.20	0.200	0.214	0.014	1120	0.0 %	1120					
2.50	2.500	2.486	-0.014	12416	0.0 %	12416					
5.00	5.000	4.947	-0.053	24648	0.0 %	24648					
7.50	7.500	7.475	-0.025	37215	0.0 %	37215					
10.0	10.000	10.048	0.048	50009	0.0 %	50009					

0.00
 1171100

LB125327

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0		211	-		0 SFAM01.1	PPB	5/8/2023 11:50	Std
0.2		1120	-		0.2 SFAM01.1	PPB	5/8/2023 11:52	Std
2.5		12416	-		2.5 SFAM01.1	PPB	5/8/2023 11:54	Std
5		24648	-		5 SFAM01.1	PPB	5/8/2023 11:57	Std
7.5		37215	-		7.5 SFAM01.1	PPB	5/8/2023 11:59	Std
10		50009	-		10 SFAM01.1	PPB	5/8/2023 12:01	Std
ICV066	ICV066	19232	3.8571 -		SFAM01.1	PPB	5/8/2023 12:04	SMPL
ICB066	ICB066	-25	-0.0166 -		SFAM01.1	PPB	5/8/2023 12:09	SMPL
CCV029	CCV029	24802	4.9776 -		SFAM01.1	PPB	5/8/2023 12:11	SMPL
CCB029	CCB029	-34	-0.0184 -		SFAM01.1	PPB	5/8/2023 12:13	SMPL
PB152633BL	PBW633	102	0.0089 -		SFAM01.1	PPB	5/8/2023 12:15	SMPL
O2215-02	MDL-WATER	341	0.057 -		SFAM01.1	PPB	5/8/2023 12:18	SMPL
CCV030	CCV0030	24915	5.0003 -		SFAM01.1	PPB	5/8/2023 12:20	SMPL
CCB030	CCB030	-10	-0.0136 -		SFAM01.1	PPB	5/8/2023 12:22	SMPL