



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB125330

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP102357
Turbidity Calibration std, 0NTU	WP102356
Turbidity Calibration std, 20NTU	WP102351
Turbidity Calibration std, 40NTU	WP102350
Turbidity Calibration std, 1NTU	WP102354
Turbidity Calibration std, 5NTU	WP102353
Turbidity Calibration - CCV std, 10 NTU	WP102352
10 NTU Standard 500 ml	W2971

Intercept: 0.0956

Slope: 0.9338

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.144	0.05		05/08/2023	09:30
2	CAL2	1	1	1.156	1.14	13.6	05/08/2023	09:34
3	CAL3	5	1	4.420	4.63	-7.4	05/08/2023	09:38
4	CAL4	10	1	9.520	10.09	0.9	05/08/2023	09:42
5	CAL5	20	1	18.900	20.14	0.7	05/08/2023	09:45
6	CAL6	40	1	37.400	39.95	-0.1	05/08/2023	09:48



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB125330

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.410	9.975	05/08/2023	09:51
2	ICB		1	0.148	0.056	05/08/2023	09:54
3	CCV1	10	1	9.340	9.900	05/08/2023	09:58
4	CCB1		1	0.150	0.058	05/08/2023	10:02
5	LB125330BL		1	0.154	0.063	05/08/2023	10:06
6	O2213-12		1	0.534	0.469	05/08/2023	10:10
7	CCV2	10	1	9.280	9.836	05/08/2023	10:13
8	CCB2		1	0.154	0.063	05/08/2023	10:17