

LB125386


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

55303



A= 0.0000e+000

B= 1.8283e-004 *slope*C= 1.9466e-003 *y-intercept*

Rho= 0.9997185

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>0/011</i>
0.00	0.000	0.013	0.013	63	0.000	63					=
0.05	0.050					302	0				2
0.20	0.200	0.204	0.004	1106	0.0 %	1106					4
2.50	2.500	2.591	0.091	14162	0.0 %	14162					-2
5.00	5.000	4.896	-0.104	26771	0.0 %	26771					-2
7.50	7.500	7.382	-0.118	40364	0.0 %	40364					1
10.0	10.000	10.113	0.113	55303	0.0 %	55303					

LB125336

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	63	-		0 SFAM01.1	PPB	5/8/2023 13:42	Std
	0.2 S01	1106	-		0.2 SFAM01.1	PPB	5/8/2023 13:45	Std
	2.5 S02	14162	-		2.5 SFAM01.1	PPB	5/8/2023 13:47	Std
	5 S03	26771	-		5 SFAM01.1	PPB	5/8/2023 13:51	Std
	7.5 S04	40364	-		7.5 SFAM01.1	PPB	5/8/2023 13:54	Std
	10 S05	55303	-		10 SFAM01.1	PPB	5/8/2023 14:03	Std
ICV067	ICV067	20877	3.8189 -		SFAM01.1	PPB	5/8/2023 14:06	SMPL
ICB067	ICB067	-129	-0.0216 -		SFAM01.1	PPB	5/8/2023 14:08	SMPL
CCV031	CCV031	27029	4.9437 -		SFAM01.1	PPB	5/8/2023 14:10	SMPL
CCB031	CCB301	-124	-0.0207 -		SFAM01.1	PPB	5/8/2023 14:12	SMPL
PB152654BL	PBS654	30	0.0074 -		SFAM01.1	PPB	5/8/2023 14:15	SMPL
O2215-04	MDL-SOIL	193	0.0372 -		SFAM01.1	PPB	5/8/2023 14:17	SMPL
CCV032	CCV032	26854	4.9117 -		SFAM01.1	PPB	5/8/2023 14:22	SMPL
CCB032	CCB032	-120	-0.02 -		SFAM01.1	PPB	5/8/2023 14:24	SMPL