


LB125342

7470A INSTRUMENT ID: CV1

 μ Abs.:

49273



Linear

A= 0.0000e+000

B= 2.0016e-004

C= -3.6699e-002

Rho= 0.9993501

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	0.042	0.042	391	0.000	391	0				
0.2	0.200	0.176	-0.024	1064	0.0 %	1064					
2.5	2.500	2.413	-0.087	12240	0.0 %	12240					
5.0	5.000	4.986	-0.014	25095	0.0 %	25095					
7.5	7.500	7.757	0.257	38935	0.0 %	38935					
10.0	10.000	9.826	-0.174	49273	0.0 %	49273					

LB125342 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 May 2023 14:58:32

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	391	-	0.000	7470A	PPB	08 May 2023 15:10:06	S	Std
0.2 - 1	50.2	1064	-	0.200	7470A	PPB	08 May 2023 15:12:24	S	Std
2.5 - 1	52.5	12240	-	2.500	7470A	PPB	08 May 2023 15:14:41	S	Std
5.0 - 1	55	25095	-	5.000	7470A	PPB	08 May 2023 15:16:58	S	Std
7.5 - 1	57.5	38935	-	7.500	7470A	PPB	08 May 2023 15:19:15	S	Std
10.0 - 1	51.5	49273	-	10.000	7470A	PPB	08 May 2023 15:21:32	S	Std
ICV68 - 1	ICV68	20901	4.1469	-	7470A	PPB	08 May 2023 15:27:12	U	SMPL
ICB68 - 1	ICB68	-216	-0.0799	-	7470A	PPB	08 May 2023 15:29:27	U	SMPL
CCV33 - 1	CCV33	24399	4.8471	-	7470A	PPB	08 May 2023 15:48:29	U	SMPL
CCB33 - 1	CCB33	-171	-0.0709	-	7470A	PPB	08 May 2023 15:50:44	U	SMPL
CRA - 1	CRA	1181	0.1997	-	7470A	PPB	08 May 2023 15:55:42	U	SMPL
HighStd - 1	HighStd	47282	9.4274	-	7470A	PPB	08 May 2023 15:57:57	U	SMPL
ChkStd - 1	ChkStd	34645	6.8979	-	7470A	PPB	08 May 2023 16:00:13	U	SMPL
PB152634BL - 1	PBW	-229	-0.0825	-	7470A	PPB	08 May 2023 16:02:31	U	SMPL
PB152634BS - 1	LCSW	17862	3.5386	-	7470A	PPB	08 May 2023 16:04:49	U	SMPL
O2611-01 - 1	357	-101	-0.0569	-	7470A	PPB	08 May 2023 16:07:05	U	SMPL
O2626-01 - 1	D2978	-15	-0.0397	-	7470A	PPB	08 May 2023 16:09:23	U	SMPL
O2630-01 - 1	2821	1011	0.1657	-	7470A	PPB	08 May 2023 16:11:40	U	SMPL
O2643-01 - 1	FSND-IDW-S-05042023	-53	-0.0473	-	7470A	PPB	08 May 2023 16:13:57	U	SMPL
O2652-02 - 1	CLAY-LINER	-34	-0.0435	-	7470A	PPB	08 May 2023 16:16:12	U	SMPL
CCV34 - 1	CCV34	24815	4.9303	-	7470A	PPB	08 May 2023 16:21:05	U	SMPL
CCB34 - 1	CCB34	-159	-0.0685	-	7470A	PPB	08 May 2023 16:23:21	U	SMPL
O2652-04 - 1	RO-290535	-37	-0.0441	-	7470A	PPB	08 May 2023 16:25:38	U	SMPL
O2652-05 - 1	COMP-DRUMS-OILY-DEBRIS	13	-0.0341	-	7470A	PPB	08 May 2023 16:27:53	U	SMPL
O2659-08 - 1	STEPS	7	-0.0353	-	7470A	PPB	08 May 2023 16:30:09	U	SMPL
O2659-09 - 1	LARGE-PAD	275	0.0183	-	7470A	PPB	08 May 2023 16:32:25	U	SMPL
O2659-10 - 1	SMALL-PAD	3461	0.6561	-	7470A	PPB	08 May 2023 16:34:41	U	SMPL
O2659-10DUP - 1	SMALL-PADDUP	3324	0.6268	-	7470A	PPB	08 May 2023 16:36:57	U	SMPL
O2659-10MS - 1	SMALL-PADMS	20280	4.0226	-	7470A	PPB	08 May 2023 16:39:14	U	SMPL
O2659-10MSD - 1	SMALL-PADMSD	21693	4.3054	-	7470A	PPB	08 May 2023 16:41:31	U	SMPL
PB152635BL - 1	PBW	-163	-0.0693	-	7470A	PPB	08 May 2023 16:43:48	U	SMPL
PB152635BS - 1	LCSW	19129	3.7922	-	7470A	PPB	08 May 2023 16:46:06	U	SMPL
CCV35 - 1	CCV35	24814	4.9301	-	7470A	PPB	08 May 2023 16:51:13	U	SMPL
CCB35 - 1	CCB35	-137	-0.0641	-	7470A	PPB	08 May 2023 16:53:29	U	SMPL
O2611-02 - 1	356	295	0.0223	-	7470A	PPB	08 May 2023 16:55:46	U	SMPL
O2611-03 - 1	358	-75	-0.0517	-	7470A	PPB	08 May 2023 16:58:01	U	SMPL
O2626-02 - 1	D2951	64	-0.0239	-	7470A	PPB	08 May 2023 17:00:17	U	SMPL
O2626-03 - 1	D2985	83	-0.0201	-	7470A	PPB	08 May 2023 17:02:34	U	SMPL
O2626-04 - 1	D3017	29	-0.0309	-	7470A	PPB	08 May 2023 17:04:50	U	SMPL
O2626-05 - 1	T2983	39	-0.0289	-	7470A	PPB	08 May 2023 17:07:07	U	SMPL
O2626-06 - 1	T2952	56	-0.0255	-	7470A	PPB	08 May 2023 17:09:25	U	SMPL
O2643-03 - 1	FSND-IDW-AQ-05042023	95	-0.0177	-	7470A	PPB	08 May 2023 17:11:41	U	SMPL
O2643-03DUP - 1	FSND-IDW-AQ-05042023DUP	99	-0.0169	-	7470A	PPB	08 May 2023 17:13:59	U	SMPL
O2643-03MS - 1	FSND-IDW-AQ-05042023MS	16675	3.3010	-	7470A	PPB	08 May 2023 17:16:15	U	SMPL
CCV36 - 1	CCV36	24035	4.7742	-	7470A	PPB	08 May 2023 17:18:30	U	SMPL
CCB36 - 1	CCB36	-122	-0.0611	-	7470A	PPB	08 May 2023 17:20:46	U	SMPL
O2643-03MSD - 1	FSND-IDW-AQ-05042023MSD	17345	3.4351	-	7470A	PPB	08 May 2023 17:25:48	U	SMPL
PB152653BL - 1	PBW	-135	-0.0637	-	7470A	PPB	08 May 2023 17:28:03	U	SMPL
PB152653BS - 1	LCSW	19355	3.8374	-	7470A	PPB	08 May 2023 17:33:09	U	SMPL
O2642-02 - 1	TT-052-IDWGW-20230504	80	-0.0207	-	7470A	PPB	08 May 2023 17:35:25	U	SMPL
O2642-02DUP - 1	TT-052-IDWGW-20230504DUP	322	0.0278	-	7470A	PPB	08 May 2023 17:37:42	U	SMPL
O2642-02MS - 1	TT-052-IDWGW-20230504MS	20715	4.1097	-	7470A	PPB	08 May 2023 17:39:58	U	SMPL
O2642-02MSD - 1	TT-052-IDWGW-20230504MSD	20197	4.0060	-	7470A	PPB	08 May 2023 17:42:15	U	SMPL
O2642-03 - 1	TT-053-IDWGW-20230504	235	0.0103	-	7470A	PPB	08 May 2023 17:44:33	U	SMPL
PB152612TB - 1	PB152612TB	72	-0.0223	-	7470A	PPB	08 May 2023 18:09:35	U	SMPL
PB152613TB - 1	PB152613TB	50	-0.0267	-	7470A	PPB	08 May 2023 18:11:52	U	SMPL
CCV38 - 1	CCV38	23443	4.6557	-	7470A	PPB	08 May 2023 18:14:10	U	SMPL
CCB38 - 1	CCB38	-62	-0.0491	-	7470A	PPB	08 May 2023 18:16:26	U	SMPL
O2659-10LX5 - 1		745	0.1124	-	7470A	PPB	08 May 2023 18:18:44	U	SMPL
O2659-10A - 1		20968	4.1603	-	7470A	PPB	08 May 2023 18:21:01	U	SMPL
O2643-03LX5 - 1		-72	-0.0511	-	7470A	PPB	08 May 2023 18:23:17	U	SMPL
O2643-03A - 1		21497	4.2662	-	7470A	PPB	08 May 2023 18:25:34	U	SMPL
O2642-02LX5 - 1		-79	-0.0525	-	7470A	PPB	08 May 2023 18:27:51	U	SMPL
O2642-02A - 1		20338	4.0342	-	7470A	PPB	08 May 2023 18:30:08	U	SMPL
CCV39 - 1	CCV39	23403	4.6477	-	7470A	PPB	08 May 2023 18:32:25	U	SMPL
CCB39 - 1	CCB39	-114	-0.0595	-	7470A	PPB	08 May 2023 18:34:42	U	SMPL

5/123