



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125345

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	05/08/2023	11:50
2	O2632-03		1	12	100.00	>212.0	>212.0	05/08/2023	12:20
3	O2632-03DUP		1	13	100.00	>212.0	>212.0	05/08/2023	12:50
4	O2648-02		1	14	100.00	>212.0	>212.0	05/08/2023	13:20
5	O2648-03		1	14	100.00	>212.0	>212.0	05/08/2023	13:50
6	O2648-04		1	12	100.00	>212.0	>212.0	05/08/2023	14:20
7	O2648-05		1	12	100.00	>212.0	>212.0	05/08/2023	14:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-5-08

WorkList ID : 169734

Department : Wet-Chemistry

Date : 05-08-2023 08:28:39

66125345

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/12/2023	Water	O2632-03	Flash Point	Cool 4 deg C	PSEG03	O11	34336	05/05/2023	1010B
05/12/2023	Water	O2648-02	Flash Point	Cool 4 deg C	PSEG03	O11	ORA-1635	05/05/2023	1010B
05/12/2023	Water	O2648-03	Flash Point	Cool 4 deg C	PSEG03	O11	ORA-1636	05/05/2023	1010B
05/12/2023	Water	O2648-04	Flash Point	Cool 4 deg C	PSEG03	O11	ORA-1637	05/05/2023	1010B
05/12/2023	Water	O2648-05	Flash Point	Cool 4 deg C	PSEG03	O11	ORA-1638	05/05/2023	1010B

Date/Time 05/08/2023 11:45
 Raw Sample Received by: RM Cwe
 Raw Sample Relinquished by: 70 Cwe

Date/Time 05/08/2023 15:30
 Raw Sample Received by: 70 Cwe
 Raw Sample Relinquished by: RM Cwe