



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125353

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	05/09/2023	11:45
2	O2658-03		1	12	100.00	>212.0	>212.0	05/09/2023	12:15
3	O2658-03DUP		1	14	100.00	>212.0	>212.0	05/09/2023	12:45
4	O2658-04		1	12	100.00	>212.0	>212.0	05/09/2023	13:15
5	O2658-05		1	12	100.00	>212.0	>212.0	05/09/2023	13:45
6	O2658-06		1	14	100.00	>212.0	>212.0	05/09/2023	14:15
7	O2658-07		1	13	100.00	>212.0	>212.0	05/09/2023	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

Lb 125358

WorkList Name : fp-5-9

WorkList ID : 169773

Department : Wet-Chemistry

Date : 05-09-2023 08:07:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/12/2023	Water	O2658-03	Flash Point	Cool 4 deg C	PSEG03	O11	327	05/05/2023	1010B
05/12/2023	Water	O2658-04	Flash Point	Cool 4 deg C	PSEG03	O11	328	05/05/2023	1010B
05/12/2023	Water	O2658-05	Flash Point	Cool 4 deg C	PSEG03	O11	507	05/05/2023	1010B
05/12/2023	Water	O2658-06	Flash Point	Cool 4 deg C	PSEG03	O11	508	05/05/2023	1010B
05/12/2023	Water	O2658-07	Flash Point	Cool 4 deg C	PSEG03	O11	509	05/05/2023	1010B

Date/Time 05/09/2023 11:35
 Raw Sample Received by: RPT
 Raw Sample Relinquished by: JPLWC1

Date/Time 05/09/2023 15:20
 Raw Sample Received by: MWC1
 Raw Sample Relinquished by: RPLWC1