

LB125355


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

49887



A= 0.0000e+000

B= 2.0166e-004

C= -1.0544e-002

Rho= 0.9998815

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/10
0.0	0.000	-0.006	-0.006	24	0.000	24	0				13
0.2	0.200	0.175	-0.025	919	0.0 %	919					4
2.5	2.500	2.596	0.096	12927	0.0 %	12927					1
5.0	5.000	4.956	-0.044	24626	0.0 %	24626					1
7.5	7.500	7.429	-0.071	36894	0.0 %	36894					1
10.0	10.000	10.050	0.050	49887	0.0 %	49887					1

LB125355 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 May 2023 10:14:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	24	-	0.0000	7470A	PPB	09 May 2023 10:38:50	S	Std
0.2 - 1	50.2	919	-	0.2000	7470A	PPB	09 May 2023 10:41:09	S	Std
2.5 - 1	52.5	12927	-	2.5000	7470A	PPB	09 May 2023 10:43:27	S	Std
5.0 - 1	55	24626	-	5.0000	7470A	PPB	09 May 2023 10:45:44	S	Std
7.5 - 1	57.5	36894	-	7.5000	7470A	PPB	09 May 2023 10:48:01	S	Std
10.0 - 1	510	49887	-	10.0000	7470A	PPB	09 May 2023 10:57:15	S	Std
ICV69 - 1	ICV69	20126	4.0480	-7470A	PPB	09 May 2023 11:00:03	U	SMPL	
ICB69 - 1	ICB69	-234	-0.0577	-7470A	PPB	09 May 2023 11:02:19	U	SMPL	
CCV39 - 1	CCV39	24431	4.9162	-7470A	PPB	09 May 2023 11:04:36	U	SMPL	
CCB39 - 1	CCB39	-222	-0.0553	-7470A	PPB	09 May 2023 11:06:51	U	SMPL	
CRA - 1	CRA	858	0.1625	-7470A	PPB	09 May 2023 11:11:40	U	SMPL	
HighStd - 1	HighStd	49193	9.9097	-7470A	PPB	09 May 2023 11:13:55	U	SMPL	
ChkStd - 1	ChkStd	33726	6.7906	-7470A	PPB	09 May 2023 11:16:10	U	SMPL	
O2648-02 - 1	ORA-1635	295	0.0489	-7470A	PPB	09 May 2023 11:18:29	U	SMPL	
O2648-05 - 1	ORA-1638	-5	-0.0116	-7470A	PPB	09 May 2023 11:20:47	U	SMPL	
O2658-03 - 1	327	2227	0.4386	-7470A	PPB	09 May 2023 11:23:05	U	SMPL	
O2658-03 - 1	328	-115	-0.0337	-7470A	PPB	09 May 2023 11:25:21	U	SMPL	
O2658-05 - 1	507	567	0.1038	-7470A	PPB	09 May 2023 11:27:40	U	SMPL	
O2658-06 - 1	508	174	0.0245	-7470A	PPB	09 May 2023 11:29:57	U	SMPL	
CCV40 - 1	CCV40	23862	4.8014	-7470A	PPB	09 May 2023 11:32:12	U	SMPL	
CCB40 - 1	CCB40	-80	-0.0267	-7470A	PPB	09 May 2023 11:34:29	U	SMPL	
O2660-01 - 1	WATER-SAMPLE-API-3-5	1840	0.3605	-7470A	PPB	09 May 2023 11:36:45	U	SMPL	
O2660-02 - 1	WATER-SAMPLE-API-3-6	1956	0.3839	-7470A	PPB	09 May 2023 11:39:01	U	SMPL	
CCV41 - 1	CCV41	23801	4.7891	-7470A	PPB	09 May 2023 11:42:57	U	SMPL	
CCB41 - 1	CCB41	-128	-0.0364	-7470A	PPB	09 May 2023 11:45:12	U	SMPL	