



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/10/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 13:50
In Date: 05/09/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:55
Out Date: 05/10/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125357

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2672-01	EYJJ9	1	1.12	8.45	9.57	8.19	83.7	
O2672-02	EYJK0	2	1.14	8.52	9.66	8.25	83.5	
O2672-03	EYJK1	3	1.17	8.38	9.55	8.11	82.8	
O2672-04	EYJK2	4	1.15	8.83	9.98	8.39	82.0	
O2672-05	EYJK3	5	1.15	8.51	9.66	8.3	84.0	
O2672-06	EYJK4	6	1.15	8.70	9.85	8.41	83.4	
O2672-07	EYJK5	7	1.13	8.85	9.98	8.51	83.4	
O2672-08	EYJK6	8	1.17	8.81	9.98	8.41	82.2	
O2672-09	EYJK7	9	1.16	8.70	9.86	8.17	80.6	
O2672-10	EYJK8	10	1.17	8.45	9.62	8.12	82.2	
O2672-11	EYJK9	11	1.18	8.75	9.93	8.21	80.3	
O2672-12	EYJL0	12	1.17	8.79	9.96	8.04	78.2	
O2672-13	EYJL1	13	1.17	8.45	9.62	8.00	80.8	
O2672-14	EYJL2	14	1.18	8.41	9.59	8.27	84.3	
O2672-15	EYJR2	15	1.15	8.83	9.98	8.29	80.9	
O2672-16	EYJR3	16	1.15	8.63	9.78	8.14	81.0	
O2672-17	EYJR4	17	1.14	8.42	9.56	8.12	82.9	
O2672-18	EYJR5	18	1.15	8.75	9.9	8.15	80.0	
O2672-19	EYJR6	19	1.17	8.45	9.62	8.1	82.0	
O2672-20	EYJR6D	20	1.17	8.45	9.62	8.1	82.0	
O2672-21	EYJR6S	21	1.17	8.45	9.62	8.1	82.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$