



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/10/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:10
In Date: 05/09/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:04
Out Date: 05/10/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125358

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2673-01	EYJL3	1	1.17	8.40	9.57	8.03	81.7	
O2673-02	EYJL4	2	1.18	8.75	9.93	8.03	78.3	
O2673-03	EYJL5	3	1.18	8.43	9.61	8.01	81.0	
O2673-04	EYJL6	4	1.17	8.81	9.98	7.64	73.4	
O2673-05	EYJL7	5	1.17	8.81	9.98	8.00	77.5	
O2673-06	EYJL8	6	1.17	8.37	9.54	7.65	77.4	
O2673-07	EYJL9	7	1.16	8.52	9.68	7.74	77.2	
O2673-08	EYJM0	8	1.18	8.44	9.62	8.33	84.7	
O2673-09	EYJM1	9	1.16	8.82	9.98	8.68	85.3	
O2673-10	EYJM2	10	1.17	8.67	9.84	8.13	80.3	
O2673-11	EYJM3	11	1.19	8.50	9.69	8.07	80.9	
O2673-12	EYJM4	12	1.19	8.54	9.73	8.24	82.6	
O2673-13	EYJM5	13	1.17	8.51	9.68	8.53	86.5	
O2673-14	EYJM6	14	1.17	8.57	9.74	7.99	79.6	
O2673-15	EYJN4	15	1.19	8.46	9.65	7.96	80.0	
O2673-16	EYJN5	16	1.19	8.79	9.98	8.37	81.7	
O2673-17	EYJN6	17	1.17	8.63	9.8	7.85	77.4	
O2673-18	EYJN7	18	1.16	8.40	9.56	8.21	83.9	
O2673-19	EYJN8	19	1.17	8.48	9.65	8.29	84.0	
O2673-20	EYJP0	20	1.17	8.48	9.65	7.77	77.8	
O2673-21	EYJP0D	21	1.17	8.48	9.65	7.77	77.8	
O2673-22	EYJP0S	22	1.17	8.48	9.65	7.77	77.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$