



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/10/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:20
In Date: 05/09/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:17
Out Date: 05/10/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125359

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2674-01	EYJM7	1	1.13	8.85	9.98	8.37	81.8	
O2674-02	EYJM8	2	1.13	8.54	9.67	8.52	86.5	
O2674-03	EYJM9	3	1.13	8.85	9.98	8.45	82.7	
O2674-04	EYJN0	4	1.14	8.52	9.66	8.24	83.3	
O2674-05	EYJN1	5	1.14	8.63	9.77	8.16	81.3	
O2674-06	EYJN2	6	1.12	8.55	9.67	7.99	80.4	
O2674-07	EYJN3	7	1.09	8.67	9.76	8.16	81.5	
O2674-08	EYJN9	8	1.09	8.48	9.57	8.25	84.4	
O2674-09	EYJP1	9	1.14	8.84	9.98	8.24	80.3	
O2674-10	EYJP2	10	1.13	8.65	9.78	8.32	83.1	
O2674-11	EYJP3	11	1.14	8.84	9.98	8.41	82.2	
O2674-12	EYJP4	12	1.13	8.57	9.7	7.89	78.9	
O2674-13	EYJP5	13	1.14	8.62	9.76	7.91	78.5	
O2674-14	EYJP6	14	1.13	8.68	9.81	8.62	86.3	
O2674-15	EYJP7	15	1.13	8.80	9.93	8.46	83.3	
O2674-16	EYJP8	16	1.13	8.72	9.85	8.11	80.0	
O2674-17	EYJP9	17	1.13	8.48	9.61	7.91	80.0	
O2674-18	EYJQ0	18	1.13	8.57	9.7	8.25	83.1	
O2674-19	EYJQ1	19	1.15	8.60	9.75	8.39	84.2	
O2674-20	EYJQ2	20	1.14	8.56	9.7	7.79	77.7	
O2674-21	EYJQ2D	21	1.14	8.56	9.7	7.79	77.7	
O2674-22	EYJQ2S	22	1.14	8.56	9.7	7.79	77.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$