

LB 125362


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

66340



A= 0.0000e+000

B= 1.5143e-004

C= 5.2566e-004

Rho= 0.9999428

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.0	0.000	0.007	0.007	43	0.000	43					
0.2	0.200	0.186	-0.014	1228	0.0 %	1228					5
2.5	2.500	2.550	0.050	16839	0.0 %	16839					2
5.0	5.000	4.968	-0.032	32801	0.0 %	32801					1
7.5	7.500	7.442	-0.058	49141	0.0 %	49141					1
10.0	10.000	10.046	0.046	66340	0.0 %	66340					0

LB125362 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 May 2023 13:48:22

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	43	-	0.00007471B	PPB		09 May 2023 14:01:49	S	Std
0.2 - 1	S0.2	1228	-	0.20007471B	PPB		09 May 2023 14:04:07	S	Std
2.5 - 1	S2.5	16839	-	2.50007471B	PPB		09 May 2023 14:06:24	S	Std
5.0 - 1	S5	32801	-	5.00007471B	PPB		09 May 2023 14:08:41	S	Std
7.5 - 1	S7.5	49141	-	7.50007471B	PPB		09 May 2023 14:10:59	S	Std
10.0 - 1	S10	66340	-	10.00007471B	PPB		09 May 2023 14:15:56	S	Std
ICV70 - 1	ICV70	26204	3.9686	-7471B	PPB		09 May 2023 14:23:38	U	SMPL
ICB70 - 1	ICB70	-101	-0.0148	-7471B	PPB		09 May 2023 14:25:53	U	SMPL
CCV42 - 1	CCV42	33092	5.0117	-7471B	PPB		09 May 2023 14:28:10	U	SMPL
CCB42 - 1	CCB42	-133	-0.0196	-7471B	PPB		09 May 2023 14:30:25	U	SMPL
CRA - 1	CRA	1183	0.1797	-7471B	PPB		09 May 2023 14:32:43	U	SMPL
HighStd - 1	HighStd	63800	9.6618	-7471B	PPB		09 May 2023 14:34:58	U	SMPL
ChkStd - 1	ChkStd	49234	7.4561	-7471B	PPB		09 May 2023 14:37:14	U	SMPL
PB152681BL - 1	PBS	-233	-0.0348	-7471B	PPB		09 May 2023 14:39:35	U	SMPL
PB152681BS - 1	LCSS	22676	3.4344	-7471B	PPB		09 May 2023 14:41:53	U	SMPL
O2646-01 - 1	CONCRETE-050523	261	0.0400	-7471B	PPB		09 May 2023 14:44:10	U	SMPL
O2652-01 - 1	CLAY-LINER	2938	0.4454	-7471B	PPB		09 May 2023 14:46:28	U	SMPL
O2658-02 - 1	501-506	67050	10.1539	-7471B	PPB		09 May 2023 14:48:46	U	SMPL
O2659-01 - 1	STEPS	6124	0.9279	-7471B	PPB		09 May 2023 14:51:04	U	SMPL
O2659-02 - 1	LARGE-PAD	123317	18.6745	-7471B	PPB		09 May 2023 14:53:22	U	SMPL
CCV43 - 1	CCV43	32190	4.8751	-7471B	PPB		09 May 2023 14:55:38	U	SMPL
CCB43 - 1	CCB43	-386	-0.0579	-7471B	PPB		09 May 2023 14:58:14	U	SMPL
O2659-03 - 1	SMALL-PAD	936927	HIGH	-7471B	PPB		09 May 2023 15:00:32	X	
O2659-04 - 1	DGA-CRUSHED-STONE	3183	0.4825	-7471B	PPB		09 May 2023 15:02:20	U	SMPL
O2665-01 - 1	TR-612-COMP-01	2078	0.3152	-7471B	PPB		09 May 2023 15:05:18	U	SMPL
O2665-08 - 1	TR-612-COMP-02	4961	0.7518	-7471B	PPB		09 May 2023 15:07:36	U	SMPL
O2666-01 - 1	BU-02-050823	4578	0.6938	-7471B	PPB		09 May 2023 15:09:53	U	SMPL
O2680-01 - 1	SU-02-050823	1437	0.2181	-7471B	PPB		09 May 2023 15:12:10	U	SMPL
O2680-01DUP - 1	SU-02-050823DUP	1571	0.2384	-7471B	PPB		09 May 2023 15:14:27	U	SMPL
O2680-01MS - 1	SU-02-050823MS	24289	3.6786	-7471B	PPB		09 May 2023 15:16:45	U	SMPL
O2680-01MSD - 1	SU-02-050823MSD	25340	3.8378	-7471B	PPB		09 May 2023 15:19:03	U	SMPL
O2681-02 - 1	VNJ-207	7021	1.0637	-7471B	PPB		09 May 2023 15:21:21	U	SMPL
CCV44 - 1	CCV44	33626	5.0925	-7471B	PPB		09 May 2023 15:23:39	U	SMPL
CCB44 - 1	CCB44	-126	-0.0186	-7471B	PPB		09 May 2023 15:25:55	U	SMPL
O2690-01 - 1	OR-3-050823	2283	0.3462	-7471B	PPB		09 May 2023 15:28:12	U	SMPL
O2691-01 - 1	AU-504-COMP-06	5829	0.8832	-7471B	PPB		09 May 2023 15:30:28	U	SMPL
O2691-08 - 1	AU-504-COMP-07	23240	3.5198	-7471B	PPB		09 May 2023 15:32:44	U	SMPL
O2680-01LX5 - 1		175	0.0270	-7471B	PPB		09 May 2023 15:35:00	U	SMPL
O2680-01A - 1		24582	3.7230	-7471B	PPB		09 May 2023 15:37:18	U	SMPL
O2658-02DLX2 - 1		34072	5.1601	-7471B	PPB		09 May 2023 15:45:38	U	SMPL
O2659-02DLX5 - 1		23728	3.5937	-7471B	PPB		09 May 2023 15:50:48	U	SMPL
O2659-03DLX10 - 1		216309	32.7563	-7471B	PPB		09 May 2023 15:53:05	U	SMPL
O2659-03DL2X50 - 1		47675	7.2200	-7471B	PPB		09 May 2023 15:57:31	U	SMPL
CCV45 - 1	CCV45	33128	5.0171	-7471B	PPB		09 May 2023 15:59:46	U	SMPL
CCB45 - 1	CCB45	-144	-0.0213	-7471B	PPB		09 May 2023 16:02:05	U	SMPL