



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/10/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 15:10
In Date: 05/09/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:33
Out Date: 05/10/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125364

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2677-01	EYJW7	1	1.13	8.71	9.84	8.01	79.0	
O2677-02	EYJW9	2	1.13	8.53	9.66	8.5	86.4	
O2677-03	EYJX0	3	1.11	8.66	9.77	8.47	85.0	
O2677-04	EYJX1	4	1.13	8.85	9.98	8.53	83.6	
O2677-05	EYJX2	5	1.11	8.79	9.9	8.36	82.5	
O2677-06	EYJX3	6	1.11	8.42	9.53	7.96	81.4	
O2677-07	EYJX4	7	1.13	8.72	9.85	8.26	81.8	
O2677-08	EYJX5	8	1.14	8.84	9.98	8.3	81.0	
O2677-09	EYJX6	9	1.12	8.69	9.81	8.14	80.8	
O2677-10	EYJX7	10	1.13	8.51	9.64	8.26	83.8	
O2677-11	EYJX8	11	1.13	8.85	9.98	8.64	84.9	
O2677-12	EYJX9	12	1.13	8.85	9.98	8.34	81.5	
O2677-13	EYJY0	13	1.1	8.83	9.93	8.19	80.3	
O2677-14	EYJY1	14	1.13	8.40	9.53	7.75	78.8	
O2677-15	EYJY2	15	1.13	8.48	9.61	7.72	77.7	
O2677-16	EYJY3	16	1.13	8.47	9.6	7.96	80.6	
O2677-17	EYJY3D	17	1.13	8.47	9.6	7.96	80.6	
O2677-18	EYJY3S	18	1.13	8.47	9.6	7.96	80.6	
O2677-19	EYJY4	19	1.13	8.68	9.81	8.12	80.5	
O2677-20	EYJY5	20	1.13	8.53	9.66	7.79	78.1	
O2677-21	EYJY6	21	1.12	8.55	9.67	7.71	77.1	
O2677-22	EYJY7	22	1.12	8.84	9.96	8.46	83.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$