



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB125365

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	05/09/2023	15:50
2	CAL2	1	Water	NA	NA	20.7	7.00	05/09/2023	15:51
3	CAL3	1	Water	NA	NA	20.7	10.02	05/09/2023	15:53
4	ICV	1	Water	NA	NA	20.6	6.98	05/09/2023	16:00
5	CCV1	1	Water	NA	NA	20.7	2.01	05/09/2023	16:02
6	O2696-01	1	Water	NA	NA	20.3	5.27	05/09/2023	16:05
7	O2696-01DUP	1	Water	NA	NA	20.4	5.28	05/09/2023	16:06
8	O2702-01	1	Water	NA	NA	22.1	5.99	05/09/2023	16:10
9	O2702-02	1	Water	NA	NA	22.1	5.91	05/09/2023	16:11
10	CCV2	1	Water	NA	NA	20.6	12.02	05/09/2023	16:15

WORKLIST(Hardcopy Internal Chain)

125365

WorkList Name : ph o2702

WorkList ID : 169799

Department : Wet-Chemistry

Date : 05-09-2023 15:32:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/08/2023	Water	O2696-01	pH	A	Cool 4 deg C	STER04	O11	POOL-WATER	
05/09/2023	Water	O2702-01	pH	C	Cool 4 deg C	PSEG03	O11	33998	05/08/2023 9040C
05/09/2023	Water	O2702-02	pH	C	Cool 4 deg C	PSEG03	O11	50033	05/09/2023 9040C
									05/09/2023 9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

15.45

70100C

ITC5M

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

05/09/23

ITC5M

40100C

18.00