



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB125366

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	05/09/2023	15:50
2	CAL2	1	Water	NA	NA	20.7	7.00	05/09/2023	15:51
3	CAL3	1	Water	NA	NA	20.7	10.02	05/09/2023	15:53
4	ICV	1	Water	NA	NA	20.6	6.98	05/09/2023	16:00
5	CCV1	1	Water	NA	NA	20.8	2.02	05/09/2023	16:20
6	O2666-01	1	Solid	20.02	20	23.1	6.84	05/09/2023	16:25
7	O2666-01DUP	1	Solid	20.03	20	23.2	6.85	05/09/2023	16:26
8	CCV2	1	Water	NA	NA	20.7	12.02	05/09/2023	16:30

WORKLIST(Hardcopy Internal Chain)

WB 125366
WB 125366

WorkList Name : ph o2666 s WorkList ID : 169800 Department : Wet-Chemistry Date : 05-09-2023 15:33:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/08/2023	Solid	02666-01	pH	A	Cool 4 deg C	PSEG03	O11	BU-02-050823	05/08/2023 9045D

Date/Time 05/09/23 15:40
Raw Sample Received by: JN(CSC)

Date/Time 05/09/23 18:00
Raw Sample Received by: JN(CSC)
Raw Sample Relinquished by: JN(CSC)