



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB125379

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	05/09/2023	18:20
2	CAL2	1	Water	NA	NA	20.7	7.00	05/09/2023	18:21
3	CAL3	1	Water	NA	NA	20.7	10.03	05/09/2023	18:22
4	ICV	1	Water	NA	NA	20.6	6.98	05/09/2023	18:25
5	CCV1	1	Water	NA	NA	20.7	2.01	05/09/2023	18:30
6	O2691-02	1	Solid	20.02	20	21.7	6.85	05/09/2023	18:33
7	O2691-02DUP	1	Solid	20.01	20	21.8	6.86	05/09/2023	18:34
8	O2691-09	1	Solid	20.02	20	21.3	5.89	05/09/2023	18:35
9	O2693-01	1	Solid	20.01	20	22.7	6.18	05/09/2023	18:36
10	CCV2	1	Water	NA	NA	20.7	12.03	05/09/2023	18:39