



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125388

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	05/10/2023	12:10
2	O2693-02		1	12	100.00	>212.0	>212.0	05/10/2023	12:40
3	O2693-02DUP		1	14	100.00	>212.0	>212.0	05/10/2023	13:10
4	O2702-01		1	13	100.00	>212.0	>212.0	05/10/2023	13:40
5	O2702-02		1	14	100.00	>212.0	>212.0	05/10/2023	14:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6125388

WorkList Name : fp-5-10

WorkList ID : 169814

Department : Wet-Chemistry

Date : 05-10-2023 08:27:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/16/2023	Water	O2693-02	Flash Point	Cool 4 deg C	PSEG03	O11	BEL-23-001	05/09/2023	1010B
05/16/2023	Water	O2702-01	Flash Point	Cool 4 deg C	PSEG03	O11	33998	05/09/2023	1010B
05/16/2023	Water	O2702-02	Flash Point	Cool 4 deg C	PSEG03	O11	50033	05/09/2023	1010B

Date/Time 05/10/2023 12:00
 Raw Sample Received by: RM CWG
 Raw Sample Relinquished by: JACOLC1

Date/Time 05/10/2023 14:50
 Raw Sample Received by: JACOLC
 Raw Sample Relinquished by: RM (CWG)