

LB125406

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

5/11/2023 11:41

Test: Phenolics-

Sample Id	^{mp/L} Result	Dil. 1 +	Response	Errors
ICV1	0.993	0.0	0.095	
ICB1	0.005	0.0	0.005	
CCV1	0.960	0.0	0.092	
CCB1	0.026	0.0	0.006	
PB152713BL	0.010	0.0	0.005	
PB152713BS	1.013	0.0	0.097	
O2618-01	-0.001	0.0	0.004	
O2618-01DUP	-0.002	0.0	0.004	
O2618-01MS	0.993	0.0	0.095	
O2618-01MSD	1.001	0.0	0.096	
O2618-03	0.004	0.0	0.004	
O2618-05	0.018	0.0	0.006	
CCV2	0.963	0.0	0.093	
CCB2	0.008	0.0	0.005	

N	14
Mean	0.428
SD	0.5029
CV%	117.48

Aquakem v. 7.2AQ1

Results from time period:

Thu May 11 09:44:31 2023

Thu May 11 11:19:38 2023

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0048	mg/l	5/11/2023 9:44:31	
0.05PPM	A	Phenolics-	P	0.0528	mg/l	5/11/2023 9:44:32	
0.1PPM	A	Phenolics-	P	0.1086	mg/l	5/11/2023 9:44:33	
0.25PPM	A	Phenolics-	P	0.2483	mg/l	5/11/2023 9:44:34	
0.50PPM	A	Phenolics-	P	0.4886	mg/l	5/11/2023 9:44:35	
1.0PPM	A	Phenolics-	P	0.9887	mg/l	5/11/2023 9:44:36	
2.0PPM	A	Phenolics-	P	2.0082	mg/l	5/11/2023 9:44:37	
ICV1	S	Phenolics-	P	0.9934	mg/l	5/11/2023 11:15:39	
ICB1	S	Phenolics-	P	0.0054	mg/l	5/11/2023 11:15:40	
CCV1	S	Phenolics-	P	0.9601	mg/l	5/11/2023 11:15:41	
CCB1	S	Phenolics-	P	0.0258	mg/l	5/11/2023 11:15:42	
PB152713BL	S	Phenolics-	P	0.0097	mg/l	5/11/2023 11:15:43	
PB152713BS	S	Phenolics-	P	1.013	mg/l	5/11/2023 11:15:44	
O2618-01	S	Phenolics-	P	-0.001	mg/l	5/11/2023 11:15:45	
O2618-01DUP	S	Phenolics-	P	-0.0018	mg/l	5/11/2023 11:15:46	
O2618-01MS	S	Phenolics-	P	0.9934	mg/l	5/11/2023 11:15:47	
O2618-01MSD	S	Phenolics-	P	1.0009	mg/l	5/11/2023 11:15:48	
O2618-03	S	Phenolics-	P	0.0042	mg/l	5/11/2023 11:15:49	
O2618-05	S	Phenolics-	P	0.0181	mg/l	5/11/2023 11:15:50	
CCV2	S	Phenolics-	P	0.9631	mg/l	5/11/2023 11:19:37	
CCB2	S	Phenolics-	P	0.0081	mg/l	5/11/2023 11:19:38	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

5/11/2023 9:58


Test Phenolics-

Accepted 5/11/2023 9:58

Factor 10.87
Bias 0.004

Coeff. of det. 0.999863

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.005	0.0048	0.0000	
2	Phenol-2	0.009	0.0528	0.0500	5.6
3	Phenol-2	0.014	0.1086	0.1000	8.6
4	Phenol-2	0.027	0.2483	0.2500	-0.7
5	Phenol-2	0.049	0.4886	0.5000	-2.3
6	Phenol-2	0.095	0.9887	1.0000	-1.1
7	Phenol-2	0.189	2.0082	2.0000	0.4

12
5/11/23