

LB125407


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

5685E



A= 0.0000e+000

B= 1.7621e-004 slope

C= -3.9707e-002 y-intercept

Rho= 0.9998877

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/O D
0.00	0.000	-0.012	-0.012	159	0.000	159					2
0.05	0.050					302	0				-24
0.20	0.200	0.153	-0.047	1091	0.0 %	1091					1
2.50	2.500	2.525	0.025	14556	0.0 %	14556					2
5.00	5.000	5.109	0.109	29222	0.0 %	29222					-1
7.50	7.500	7.446	-0.054	42483	0.0 %	42483					0
10.0	10.000	9.979	-0.021	56856	0.0 %	56856					

LB125407

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	159	-		0 SFAM01.1	PPB	5/11/2023 12:13	Std
	0.2 S01	1091	-		0.2 SFAM01.1	PPB	5/11/2023 12:16	Std
	2.5 S02	14556	-		2.5 SFAM01.1	PPB	5/11/2023 12:18	Std
	5 S03	29222	-		5 SFAM01.1	PPB	5/11/2023 12:23	Std
	7.5 S04	42483	-		7.5 SFAM01.1	PPB	5/11/2023 12:25	Std
	10 S05	56856	-		10 SFAM01.1	PPB	5/11/2023 12:32	Std
ICV73	ICV73	21686	3.7815 -		SFAM01.1	PPB	5/11/2023 12:35	SMPL
ICB73	ICB73	46	-0.0316 -		SFAM01.1	PPB	5/11/2023 12:37	SMPL
CCV49	CCV49	29038	5.0769 -		SFAM01.1	PPB	5/11/2023 12:39	SMPL
CCB49	CCB49	17	-0.0367 -		SFAM01.1	PPB	5/11/2023 12:41	SMPL
PB152695BL	PB152695BL	410	0.0325 -		SFAM01.1	PPB	5/11/2023 12:44	SMPL
O2671-01	YBRR0	103361	18.1731 -		SFAM01.1	PPB	5/11/2023 12:46	SMPL
O2671-02	YBRR0D	100653	17.6959 -		SFAM01.1	PPB	5/11/2023 12:48	SMPL
O2671-03	YBRR0S	112875	19.8495 -		SFAM01.1	PPB	5/11/2023 12:50	SMPL
O2671-04	YBRR1	300179	52.8535 -		SFAM01.1	PPB	5/11/2023 12:53	SMPL
O2671-05	YBRR2	93796	16.4877 -		SFAM01.1	PPB	5/11/2023 12:55	SMPL
O2671-06	YBRR3	191147	33.6414 -		SFAM01.1	PPB	5/11/2023 12:58	SMPL
O2671-07	YBRR4	23082	4.0275 -		SFAM01.1	PPB	5/11/2023 13:00	SMPL
O2671-08	YBRR5	3852	0.639 -		SFAM01.1	PPB	5/11/2023 13:02	SMPL
O2671-09	YBRR6	3559	0.5874 -		SFAM01.1	PPB	5/11/2023 13:05	SMPL
O2671-01DLX5	YBRR0	21278	3.7096 -		SFAM01.1	PPB	5/11/2023 13:10	SMPL
O2671-02DLX5	YBRR0D	21487	3.7464 -		SFAM01.1	PPB	5/11/2023 13:12	SMPL
O2671-03DLX5	YBRR0S	23590	4.117 -		SFAM01.1	PPB	5/11/2023 13:14	SMPL
O2671-04DLX10	YBRR1	32798	5.7395 -		SFAM01.1	PPB	5/11/2023 13:17	SMPL
O2671-05DLX5	YBRR2	19584	3.4111 -		SFAM01.1	PPB	5/11/2023 13:19	SMPL
O2671-06DLX10	YBRR3	19866	3.4608 -		SFAM01.1	PPB	5/11/2023 13:21	SMPL
CCV50	CCV50	29765	5.205 -		SFAM01.1	PPB	5/11/2023 13:23	SMPL
CCB50	CCB50	40	-0.0327 -		SFAM01.1	PPB	5/11/2023 13:26	SMPL