



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB125408

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	05/11/2023	11:50
2	CAL2	1	Water	NA	NA	20.6	7.00	05/11/2023	11:51
3	CAL3	1	Water	NA	NA	20.6	10.02	05/11/2023	11:52
4	ICV	1	Water	NA	NA	20.6	6.98	05/11/2023	12:00
5	CCV1	1	Water	NA	NA	20.7	2.01	05/11/2023	12:03
6	O2755-01	1	Water	NA	NA	20.2	6.78	05/11/2023	12:10
7	O2755-01DUP	1	Water	NA	NA	20.2	6.79	05/11/2023	12:11
8	CCV2	1	Water	NA	NA	20.6	12.02	05/11/2023	12:15

VB 125408

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 02755

WorkList ID : 169899

Department : Wet-Chemistry

Date : 05-11-2023 11:39:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/11/2023	Water	02755-01	pH C	Cool 4 deg C	PSEG03	O11	A5560	05/11/2023	9040C

Date/Time 05-11-23 11:45
Raw Sample Received by: 706001
Raw Sample Relinquished by: 12604

Date/Time 05-11-23 13:00
Raw Sample Received by: 12604
Raw Sample Relinquished by: 706001