



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/12/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 13:55
In Date: 05/11/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 05/12/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125409

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2698-01	EYKE5	1	1.14	8.85	9.99	8.45	82.6	
O2698-02	EYKE6	2	1.16	8.82	9.98	8.92	88.0	
O2698-03	EYKE7	3	1.18	8.79	9.97	8.41	82.3	
O2698-04	EYKE8	4	1.18	8.42	9.6	8.22	83.6	
O2698-05	EYKE9	5	1.19	8.60	9.79	8.32	82.9	
O2698-06	EYKF0	6	1.19	8.33	9.52	7.74	78.6	
O2698-07	EYKF1	7	1.16	8.72	9.88	7.77	75.8	
O2698-08	EYKF2	8	1.19	8.78	9.97	7.84	75.7	
O2698-09	EYKF3	9	1.19	8.50	9.69	8.18	82.2	
O2698-10	EYKF4	10	1.19	8.69	9.88	8.04	78.8	
O2698-11	EYKF5	11	1.19	8.62	9.81	7.65	74.9	
O2698-12	EYKF6	12	1.12	8.51	9.63	7.62	76.4	
O2698-13	EYKF7	13	1.19	8.56	9.75	8.39	84.1	
O2698-14	EYKF8	14	1.19	8.76	9.95	8.29	81.1	
O2698-15	EYKF9	15	1.17	8.51	9.68	7.96	79.8	
O2698-16	EYKG0	16	1.19	8.61	9.8	8.03	79.4	
O2698-17	EYKG1	17	1.13	8.48	9.61	8.23	83.7	
O2698-18	EYKG2	18	1.16	8.51	9.67	8.45	85.7	
O2698-19	EYKG3	19	1.19	8.58	9.77	8.28	82.6	
O2698-20	EYKG4	20	1.15	8.50	9.65	7.89	79.3	
O2698-21	EYKG4D	21	1.15	8.50	9.65	7.89	79.3	
O2698-22	EYKG4S	22	1.15	8.50	9.65	7.89	79.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$