



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/12/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:20
In Date: 05/11/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:25
Out Date: 05/12/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125411

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2701-01	EYKJ5	1	1.13	8.46	9.59	7.11	70.7	
O2701-02	EYKJ6	2	1.19	8.73	9.92	8.09	79.0	
O2701-03	EYKJ7	3	1.13	8.80	9.93	8.15	79.8	
O2701-04	EYKJ7D	4	1.13	8.80	9.93	8.15	79.8	
O2701-05	EYKJ7S	5	1.13	8.80	9.93	8.15	79.8	
O2701-06	EYKJ8	6	1.18	8.70	9.88	8.13	79.9	
O2701-07	EYKJ9	7	1.19	8.39	9.58	8.22	83.8	
O2701-08	EYKK0	8	1.18	8.58	9.76	8.00	79.5	
O2701-09	EYKK1	9	1.16	8.44	9.6	7.97	80.7	
O2701-10	EYKK2	10	1.1	8.47	9.57	7.9	80.3	
O2701-11	EYKK3	11	1.19	8.49	9.68	8.31	83.9	
O2701-12	EYKK4	12	1.19	8.73	9.92	8.34	81.9	
O2701-13	EYKK5	13	1.19	8.55	9.74	7.99	79.5	
O2701-14	EYKK6	14	1.15	8.35	9.5	7.89	80.7	
O2701-15	EYKK7	15	1.19	8.36	9.55	7.84	79.5	
O2701-16	EYKK8	16	1.19	8.43	9.62	8.32	84.6	
O2701-17	EYKK9	17	1.19	8.58	9.77	8.32	83.1	
O2701-18	EYKL0	18	1.14	8.82	9.96	8.57	84.2	
O2701-19	EYKL1	19	1.16	8.80	9.96	8.28	80.9	
O2701-20	EYKL2	20	1.11	8.46	9.57	8.28	84.8	
O2701-21	EYKL3	21	1.14	8.47	9.61	8.03	81.3	
O2701-22	EYKL4	22	1.1	8.68	9.78	8.02	79.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$