



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/12/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:35
In Date: 05/11/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 05/12/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125412

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2703-01	EYKL5	1	1.18	8.52	9.7	7.96	79.6	
O2703-02	EYKL6	2	1.19	8.67	9.86	8.46	83.9	
O2703-03	EYKL7	3	1.19	8.61	9.8	8.2	81.4	
O2703-04	EYKL8	4	1.15	8.42	9.57	8.21	83.8	
O2703-05	EYKL9	5	1.15	8.39	9.54	8.07	82.5	
O2703-06	EYKM0	6	1.19	8.71	9.9	8.36	82.3	
O2703-07	EYKM1	7	1.16	8.81	9.97	8.82	86.9	
O2703-08	EYKM2	8	1.17	8.60	9.77	8.22	82.0	
O2703-09	EYKM3	9	1.16	8.81	9.97	9.3	92.4	
O2703-10	EYKM3D	10	1.16	8.81	9.97	9.3	92.4	
O2703-11	EYKM3S	11	1.16	8.81	9.97	9.3	92.4	
O2703-12	EYKM4	12	1.19	8.33	9.52	8.06	82.5	
O2703-13	EYKM5	13	1.19	8.40	9.59	8.01	81.2	
O2703-14	EYKM6	14	1.18	8.79	9.97	8.27	80.7	
O2703-15	EYKM7	15	1.12	8.60	9.72	8.46	85.3	
O2703-16	EYKM8	16	1.16	8.66	9.82	8.48	84.5	
O2703-17	EYKM9	17	1.14	8.43	9.57	8.32	85.2	
O2703-18	EYKN0	18	1.19	8.75	9.94	8.52	83.8	
O2703-19	EYKN1	19	1.19	8.39	9.58	7.64	76.9	
O2703-20	EYKN2	20	1.13	8.81	9.94	7.84	76.2	
O2703-21	EYKN3	21	1.19	8.58	9.77	7.9	78.2	
O2703-22	EYKN4	22	1.18	8.71	9.89	8.33	82.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$